



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB101027

BalanceID: WC SC-4

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP71996
COD calibration std. 100 ppm	WP71994
COD calibration std. 10 ppm	WP71992
COD calibration std. 150 ppm	WP71995
COD calibration std. 50 ppm	WP71993
COD calibration std. 0 ppm	WP71990
COD CCV std, 50ppm	WP71997
COD LOD std, 5ppm	WP71998
COD LOQ std, 10.0PPM	WP71999
COD Digestion Vials Low Level 0-150Mg/L	W2468

Temp In (C): 150	Date In: 02/26/2019	Time In: 11:00
Temp Out (C): 150	Date Out: 02/26/2019	Time Out: 13:01

Intercept: 0.2078

Slope: 0.9999

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	02/26/2019	14:00
2	CAL2	10	1	Water	11.000	02/26/2019	14:00
3	CAL3	50	1	Water	50.000	02/26/2019	14:01
4	CAL4	100	1	Water	99.000	02/26/2019	14:01
5	CAL5	150	1	Water	151.000	02/26/2019	14:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB101027

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.797	02/26/2019	14:02
2	ICB				1	Water	0.000	-0.208	02/26/2019	14:03
3	CCV1	50			1	Water	50.000	49.797	02/26/2019	14:14
4	CCB1				1	Water	0.000	-0.208	02/26/2019	14:15
5	LB101027BL		1.00	50	1	Solid	0.000	-0.208	02/26/2019	14:15
6	LB101027BS	50	1.01	50	1	Solid	50.000	49.797	02/26/2019	14:16
7	K1560-01		1.02	50	1	Solid	6.000	5.793	02/26/2019	14:16
8	K1560-02		1.01	50	1	Solid	13.000	12.793	02/26/2019	14:17
9	CCV2	50			1	Water	50.000	49.797	02/26/2019	14:17
10	CCB2				1	Water	0.000	-0.208	02/26/2019	14:18

WORKLIST(Hardcopy Internal Chain)

LB101027

WorkList Name : CODs022619

WorkList ID : 122717

Date : 02-26-2019 10:22:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2019	Solid	K1560-01	COD	Cool 4 deg C	CHEM02	Wet Chem L	LOD-MDL-SOIL-01-QT1-2019	02/21/2019	SM5220 D
03/12/2019	Solid	K1560-02	COD	Cool 4 deg C	CHEM02	Wet Chem L	LOQ-SOIL-02-QT1-2019	02/21/2019	SM5220 D

Date/Time _____
 Received by: _____
 Relinquished by: _____

N/A

Date/Time _____
 Received by: _____
 Relinquished by: _____

N/A