

LB101029


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47926



A= 0.0000e+000

B= 2.0987e-004

C= -1.4878e-002

Rho= 0.9999290

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/h D
0.00	0.000	-0.008	-0.008	35	0.000	35					1
0.2	0.200	0.202	0.002	1034	0.0 %	1034					3
2.5	2.500	2.564	0.064	12289	0.0 %	12289					-1
5.0	5.000	4.936	-0.064	23590	0.0 %	23590					-1
7.5	7.500	7.462	-0.038	35627	0.0 %	35627					0
10	10.000	10.043	0.043	47926	0.0 %	47926					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	35	-		0 ISM02.4	PPB	2/26/2019 17:31	Std
	0.2 S0.2	1034	-		0.2 ISM02.4	PPB	2/26/2019 17:34	Std
	2.5 S2.5	12289	-		2.5 ISM02.4	PPB	2/26/2019 17:36	Std
	5 S5.0	23590	-		5 ISM02.4	PPB	2/26/2019 17:38	Std
	7.5 S7.5	35627	-		7.5 ISM02.4	PPB	2/26/2019 17:40	Std
	10 S10.0	47926	-		10 ISM02.4	PPB	2/26/2019 17:43	Std
ICV	ICV	18890	3.9495	-	ISM02.4	PPB	2/26/2019 17:46	SMPL
ICB	ICB	4	-0.014	-	ISM02.4	PPB	2/26/2019 17:48	SMPL
CCV027	CCV027	24086	5.04	-	ISM02.4	PPB	2/26/2019 17:50	SMPL
CCB027	CCB027	-64	-0.0283	-	ISM02.4	PPB	2/26/2019 17:52	SMPL
PB117379BL	PBW379	11	-0.0126	-	ISM02.4	PPB	2/26/2019 17:55	SMPL
K1584-01	MCOAN9	25	-0.0096	-	ISM02.4	PPB	2/26/2019 17:57	SMPL
K1584-02	MCOAPO	10	-0.0128	-	ISM02.4	PPB	2/26/2019 17:59	SMPL
K1584-03	MCOAPOD	21	-0.0105	-	ISM02.4	PPB	2/26/2019 18:01	SMPL
K1584-04	MCOAPOS	4749	0.9818	-	ISM02.4	PPB	2/26/2019 18:04	SMPL
PB117380BL	PBW380	1	-0.0147	-	ISM02.4	PPB	2/26/2019 18:06	SMPL
K1523-01	MESXG4	3	-0.0142	-	ISM02.4	PPB	2/26/2019 18:08	SMPL
K1523-02	MESXG4D	27	-0.0092	-	ISM02.4	PPB	2/26/2019 18:10	SMPL
K1523-03	MESXG4S	4839	1.0007	-	ISM02.4	PPB	2/26/2019 18:13	SMPL
K1523-04	MESXG5	20	-0.0107	-	ISM02.4	PPB	2/26/2019 18:15	SMPL
PB117381BL	PBW738	20	-0.0107	-	ISM02.4	PPB	2/26/2019 18:17	SMPL
K1555-12	MBF5X9	27	-0.0092	-	ISM02.4	PPB	2/26/2019 18:19	SMPL
K1555-13	MBF5Y2	200	0.0271	-	ISM02.4	PPB	2/26/2019 18:22	SMPL
K1555-14	MBF5Y3	166	0.02	-	ISM02.4	PPB	2/26/2019 18:24	SMPL
K1555-15	MBF5Y4	240	0.0355	-	ISM02.4	PPB	2/26/2019 18:26	SMPL
K1555-16	MBF5Y5	138	0.0141	-	ISM02.4	PPB	2/26/2019 18:29	SMPL
K1555-17	MBF5Y5D	157	0.0181	-	ISM02.4	PPB	2/26/2019 18:31	SMPL
K1555-18	MBF5Y5S	5137	1.0632	-	ISM02.4	PPB	2/26/2019 18:33	SMPL
K1555-19	MBF5Y6	152	0.017	-	ISM02.4	PPB	2/26/2019 18:35	SMPL
K1555-20	MBF5Y7	149	0.0164	-	ISM02.4	PPB	2/26/2019 18:38	SMPL
CCV028	CCV028	24973	5.2261	-	ISM02.4	PPB	2/26/2019 18:40	SMPL
CCB028	CCB028	-33	-0.0218	-	ISM02.4	PPB	2/26/2019 18:42	SMPL
K1555-21	MBF5Y8	118	0.0099	-	ISM02.4	PPB	2/26/2019 18:44	SMPL
K1555-22	MBF5Y9	122	0.0107	-	ISM02.4	PPB	2/26/2019 18:47	SMPL
K1555-23	MBF5Z0	139	0.0143	-	ISM02.4	PPB	2/26/2019 18:49	SMPL
CCV029	CCV029	24411	5.1082	-	ISM02.4	PPB	2/26/2019 18:51	SMPL
CCB029	CCB029	-53	-0.026	-	ISM02.4	PPB	2/26/2019 18:53	SMPL