


LB101032

INSTRUMENT ID : CV1

ISM02.4



Linear

A= 0.0000e+000

B= 1.9558e-004

C= 1.1267e-002

Rho= 0.9991413

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.022	0.022	57	0.000	57					
0.2	0.200	0.210	0.010	1018	0.0 %	1018					5
2.5	2.500	2.585	0.085	13161	0.0 %	13161					3
5.0	5.000	4.722	-0.278	24087	0.0 %	24087					-6
7.5	7.500	7.726	0.226	39443	0.0 %	39443					3
10	10.000	9.934	-0.066	50736	0.0 %	50736					-1

LB101032
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	57	-		0 ISM02.4	PPB	2/27/2019 8:52	Std
	0.2 S0.2	1018	-		0.2 ISM02.4	PPB	2/27/2019 8:55	Std
	2.5 S2.5	13161	-		2.5 ISM02.4	PPB	2/27/2019 8:57	Std
	5 S5.0	24087	-		5 ISM02.4	PPB	2/27/2019 8:59	Std
	7.5 S7.5	39443	-		7.5 ISM02.4	PPB	2/27/2019 9:01	Std
	10 S10.0	50736	-		10 ISM02.4	PPB	2/27/2019 9:04	Std
ICV	ICV	20236	3.969 -		ISM02.4	PPB	2/27/2019 9:07	SMPL
ICB	ICB	11	0.0134 -		ISM02.4	PPB	2/27/2019 9:10	SMPL
CCV030	CCV030	24215	4.7472 -		ISM02.4	PPB	2/27/2019 9:12	SMPL
CCB030	CCB030	-11	0.0091 -		ISM02.4	PPB	2/27/2019 9:14	SMPL
PB117377BL	PBS377	50	0.021 -		ISM02.4	PPB	2/27/2019 9:20	SMPL
K1555-01	MBF5W8	1369	0.279 -		ISM02.4	PPB	2/27/2019 9:22	SMPL
K1555-02	MBF5W9	917	0.1906 -		ISM02.4	PPB	2/27/2019 9:24	SMPL
K1555-03	MBF5X0	931	0.1934 -		ISM02.4	PPB	2/27/2019 9:26	SMPL
K1555-04	MBF5X1	843	0.1761 -		ISM02.4	PPB	2/27/2019 9:29	SMPL
K1555-05	MBF5X1D	857	0.1789 -		ISM02.4	PPB	2/27/2019 9:31	SMPL
K1555-06	MBF5X1S	15591	3.0606 -		ISM02.4	PPB	2/27/2019 9:37	SMPL
K1555-07	MBF5X2	1208	0.2475 -		ISM02.4	PPB	2/27/2019 9:39	SMPL
K1555-08	MBF5X3	2301	0.4613 -		ISM02.4	PPB	2/27/2019 9:42	SMPL
K1555-09	MBF5X4	2733	0.5458 -		ISM02.4	PPB	2/27/2019 9:44	SMPL
K1555-10	MBF5X5	1020	0.2108 -		ISM02.4	PPB	2/27/2019 9:46	SMPL
K1555-11	MBF5X7	2484	0.4971 -		ISM02.4	PPB	2/27/2019 9:48	SMPL
CCV031	CCV031	24100	4.7247 -		ISM02.4	PPB	2/27/2019 9:51	SMPL
CCB031	CCB031	-47	0.0021 -		ISM02.4	PPB	2/27/2019 9:53	SMPL