



Analytical Summary Report

Analysis Method: 9041A

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101035

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	02/26/2019	16:35
2	CAL2	1	Water	NA	NA	22.3	7.01	02/26/2019	16:36
3	CAL3	1	Water	NA	NA	22.3	10.02	02/26/2019	16:37
4	ICV	1	Water	NA	NA	22.2	6.99	02/26/2019	16:38
5	CCV1	1	Water	NA	NA	22.3	2.01	02/26/2019	16:39
6	K1632-01	1	Solid	20.02	20	23.1	5.78	02/26/2019	16:40
7	K1632-01DUP	1	Solid	20.03	20	23.2	5.77	02/26/2019	16:41
8	CCV2	1	Water	NA	NA	22.2	12.02	02/26/2019	16:42

18101035

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Phs022619 WorkList ID : 122744 Date : 02-26-2019 16:25:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2019	Solid	K1632-01	pH	Cool 4 deg C	PSEG01	K31	50153	02/26/2019	9041A

Date/Time 02-26-19 1630
Received by: J1
Relinquished by: CD

Date/Time 02-26-19 17:00
Received by: JP
Relinquished by: SA