



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101036

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	02/26/2019	16:35
2	CAL2	1	Water	NA	NA	22.3	7.01	02/26/2019	16:36
3	CAL3	1	Water	NA	NA	22.3	10.02	02/26/2019	16:37
4	ICV	1	Water	NA	NA	22.2	6.99	02/26/2019	16:38
5	CCV1	1	Water	NA	NA	22.1	2.02	02/26/2019	18:10
6	K1350-12	1	Solid	20.02	20	21.9	7.65	02/26/2019	18:11
7	K1350-12DUP	1	Solid	20.03	20	21.8	7.66	02/26/2019	18:12
8	K1350-14	1	Solid	20.04	20	21.8	8.05	02/26/2019	18:13
9	K1350-16	1	Solid	20.02	20	21.9	8.45	02/26/2019	18:15
10	K1353-02	1	Solid	20.03	20	21.8	8.60	02/26/2019	18:16
11	K1353-04	1	Solid	20.04	20	21.7	7.85	02/26/2019	18:18
12	K1353-06	1	Solid	20.02	20	21.5	7.17	02/26/2019	18:20
13	K1353-08	1	Solid	20.03	20	21.3	8.26	02/26/2019	18:21
14	K1619-01	1	Solid	20.04	20	21.0	7.25	02/26/2019	18:22
15	K1633-04	1	Solid	20.03	20	22.1	6.95	02/26/2019	18:24
16	CCV2	1	Water	NA	NA	22.2	12.02	02/26/2019	18:25
17	K1633-08	1	Solid	20.04	20	21.6	5.98	02/26/2019	18:26
18	CCV3	1	Water	NA	NA	22.1	2.02	02/26/2019	18:27

WORKLIST(Hardcopy Internal Chain)

LB 101036

WorkList Name : CORROSIONITY K1606

WorkList ID : 122762

Date : 02-27-2019 08:17:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2019	Solid	K1350-12	Corrosivity	Cool 4 deg C	PSEG05		EO-01-0222619-A	02/26/2019	9045D
02/26/2019	Solid	K1350-14	Corrosivity	Cool 4 deg C	PSEG05		EO-01-0222619-B	02/26/2019	9045D
02/26/2019	Solid	K1350-16	Corrosivity	Cool 4 deg C	PSEG05		EO-01-0222619-C	02/26/2019	9045D
02/26/2019	Solid	K1353-02	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-02-0222619-A	02/26/2019	9045D
02/26/2019	Solid	K1353-04	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-02-0222619-B	02/26/2019	9045D
02/26/2019	Solid	K1353-06	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-02-0222619-C	02/26/2019	9045D
02/26/2019	Solid	K1353-08	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-02-0222619-D	02/26/2019	9045D
02/20/2019	Solid	K1619-01	Corrosivity	Cool 4 deg C	CLEA11	P51	OIL-WATER-SEPERATOR-2	02/20/2019	9045D
02/26/2019	Solid	K1633-04	Corrosivity	Cool 4 deg C	PSEG01	K31	SR-11-S	02/26/2019	9045D
02/26/2019	Solid	K1633-08	Corrosivity	Cool 4 deg C	PSEG01	K31	SR-11-D	02/26/2019	9045D

Date/Time 02-26-19 18:00

Received by: sn

Relinquished by: CB

Date/Time 02-26-19 18:40

Received by: sn

Relinquished by: sn