



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101038

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/26/2019	13:25
2	CAL2	1	Water	NA	NA	22.1	7.00	02/26/2019	13:26
3	CAL3	1	Water	NA	NA	22.2	10.02	02/26/2019	13:27
4	ICV	1	Water	NA	NA	22.2	7.01	02/26/2019	13:28
5	CCV1	1	Water	NA	NA	22.2	2.01	02/26/2019	15:50
6	K1631-14	1	Water	NA	NA	20.6	7.11	02/26/2019	16:00
7	K1631-14DUP	1	Water	NA	NA	20.7	7.12	02/26/2019	16:01
8	CCV2	1	Water	NA	NA	22.2	12.02	02/26/2019	16:04

LB 101038

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K1614

WorkList ID : 122763

Date : 02-27-2019 08:18:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2019	Water	K1631-14	Corrosivity	Cool 4 deg C	PSEG01	K31	Z31W	02/26/2019	9040C

Date/Time 02-26-19 13:20
Received by: SP
Relinquished by: CQ

Date/Time 02-26-19 15:00
Received by: SP
Relinquished by: CQ