

=====

Test results

Aquakem 7.2AQ1

2B101045
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

2/27/2019 13:17

Test: Total CN

Sample Id	Result <i>Mg/d</i>	Dil. 1 +	Response	Errors
ICV1	94.904	0.0	0.092	
ICB1	-0.280	0.0	0.004	
CCV1	234.863	0.0	0.221	
CCB1	-0.458	0.0	0.004	
PB117320BL	-0.478	0.0	0.004	
PB117320BS	95.508	0.0	0.092	
MIDPB117320	234.501	0.0	0.220	
K1560-12	2.894	0.0	0.007	
CCV2	234.285	0.0	0.220	
CCB2	-0.288	0.0	0.004	

N	10
Mean	89.545
SD	106.9761
CV%	119.47

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 27 13:14:24 2019

Wed Feb 27 13:14:34 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.5595	µg/l	2/27/2019 12:02:31	
5.0PPBCN	A	Total CN	P	4.7574	µg/l	2/27/2019 12:02:32	
10PPBCN	A	Total CN	P	9.8444	µg/l	2/27/2019 12:02:33	
50PPBCN	A	Total CN	P	49.917	µg/l	2/27/2019 12:02:34	
100PPBCN	A	Total CN	P	101.1622	µg/l	2/27/2019 12:02:35	
250PPBCN	A	Total CN	P	250.1942	µg/l	2/27/2019 12:02:36	
500PPBCN	A	Total CN	P	499.6843	µg/l	2/27/2019 12:02:37	
ICV1	S	Total CN	P	94.9044	µg/l	2/27/2019 13:14:24	
ICB1	S	Total CN	P	-0.2804	µg/l	2/27/2019 13:14:25	
CCV1	S	Total CN	P	234.8626	µg/l	2/27/2019 13:14:26	
CCB1	S	Total CN	P	-0.4581	µg/l	2/27/2019 13:14:27	
PB117320BL	S	Total CN	P	-0.4782	µg/l	2/27/2019 13:14:28	
PB117320BS	S	Total CN	P	95.5076	µg/l	2/27/2019 13:14:29	
MIDPB117320	S	Total CN	P	234.5005	µg/l	2/27/2019 13:14:30	
K1560-12	S	Total CN	P	2.8936	µg/l	2/27/2019 13:14:32	
CCV2	S	Total CN	P	234.2849	µg/l	2/27/2019 13:14:33	
CCB2	S	Total CN	P	-0.2875	µg/l	2/27/2019 13:14:34	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

2/27/2019 12:04


Test Total CN

Accepted 2/27/2019 12:04

Factor 1086
Bias 0.004

Coeff. of det. 0.999991

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.5595	0.0000	
2	5.0PPBCN	0.009	4.7574	5.0000	
3	10PPBCN	0.013	9.8444	10.0000	
4	50PPBCN	0.050	49.9170	50.0000	
5	100PPBCN	0.097	101.1622	100.0000	
6	250PPBCN	0.235	250.1942	250.0000	
7	500PPBCN	0.464	499.6843	500.0000	