

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

2/27/2019 14:33

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	1.000	0.0	0.121	
ICB1	0.007	0.0	0.015	
CCV1	0.989	0.0	0.120	
CCB1	0.005	0.0	0.015	
PB117331BL	0.006	0.0	0.015	
PB117331BS	1.005	0.0	0.122	
K1570-01	0.013	0.0	0.016	
K1570-01DUP	0.008	0.0	0.015	
K1570-01MS	0.913	0.0	0.112	
K1570-01MSD	0.919	0.0	0.112	
K1570-02	0.022	0.0	0.016	
K1570-03	0.016	0.0	0.016	
K1570-04	0.014	0.0	0.016	
K1570-05	0.030	0.0	0.017	
CCV2	0.973	0.0	0.118	
CCB2	0.007	0.0	0.015	
K1587-01	0.003	0.0	0.014	
K1560-10	0.015	0.0	0.016	
K1560-11	0.102	0.0	0.025	
CCV3	0.981	0.0	0.119	
CCB3	0.008	0.0	0.015	

N	21
Mean	0.335
SD	0.4600
CV%	137.28

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 27 13:47:03 2019

Wed Feb 27 14:29:56 2019

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0477	mg/l	2/27/2019 13:47:03	
0.1PPM	A	Phenolics-	P	0.1048	mg/l	2/27/2019 13:47:04	
0.25PPM	A	Phenolics-	P	0.2561	mg/l	2/27/2019 13:47:05	
0.50PPM	A	Phenolics-	P	0.4981	mg/l	2/27/2019 13:47:06	
1.0PPM	A	Phenolics-	P	0.9876	mg/l	2/27/2019 13:47:07	
2.0PPM	A	Phenolics-	P	2.0057	mg/l	2/27/2019 13:47:08	
ICV1	S	Phenolics-	P	1.0003	mg/l	2/27/2019 14:14:57	
ICB1	S	Phenolics-	P	0.0066	mg/l	2/27/2019 14:14:58	
CCV1	S	Phenolics-	P	0.9894	mg/l	2/27/2019 14:14:59	
CCB1	S	Phenolics-	P	0.0054	mg/l	2/27/2019 14:15:00	
PB117331BL	S	Phenolics-	P	0.0064	mg/l	2/27/2019 14:15:01	
PB117331BS	S	Phenolics-	P	1.0049	mg/l	2/27/2019 14:15:02	
K1570-01	S	Phenolics-	P	0.0133	mg/l	2/27/2019 14:15:03	
K1570-01DUP	S	Phenolics-	P	0.0077	mg/l	2/27/2019 14:15:04	
K1570-01MS	S	Phenolics-	P	0.9134	mg/l	2/27/2019 14:15:05	
K1570-01MSD	S	Phenolics-	P	0.9192	mg/l	2/27/2019 14:15:06	
K1570-02	S	Phenolics-	P	0.0216	mg/l	2/27/2019 14:15:07	
K1570-03	S	Phenolics-	P	0.0156	mg/l	2/27/2019 14:15:08	
K1570-04	S	Phenolics-	P	0.0137	mg/l	2/27/2019 14:24:59	
K1570-05	S	Phenolics-	P	0.03	mg/l	2/27/2019 14:25:00	
CCV2	S	Phenolics-	P	0.9733	mg/l	2/27/2019 14:25:01	
CCB2	S	Phenolics-	P	0.0068	mg/l	2/27/2019 14:25:02	
K1587-01	S	Phenolics-	P	0.0031	mg/l	2/27/2019 14:25:03	
K1560-10	S	Phenolics-	P	0.0155	mg/l	2/27/2019 14:25:04	
K1560-11	S	Phenolics-	P	0.1021	mg/l	2/27/2019 14:25:04	
CCV3	S	Phenolics-	P	0.9806	mg/l	2/27/2019 14:29:54	
CCB3	S	Phenolics-	P	0.0079	mg/l	2/27/2019 14:29:55	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

2/27/2019 13:48


Test Phenolics-

Accepted 2/27/2019 13:48

Factor 9.35
Bias 0.014

Coeff. of det. 0.999908

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.019	0.0477	0.0500	
2	Phenol-2	0.025	0.1048	0.1000	
3	Phenol-2	0.042	0.2561	0.2500	
4	Phenol-2	0.067	0.4981	0.5000	
5	Phenol-2	0.120	0.9876	1.0000	
6	Phenol-2	0.229	2.0057	2.0000	