

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

2/27/2019 15:06

Test: Phenolics-

Sample Id	Result ^{mg/d}	Dil. 1 +	Response	Errors
ICV1	1.002	0.0	0.121	
ICB1	0.002	0.0	0.014	
CCV1	0.975	0.0	0.118	
CCB1	0.005	0.0	0.015	
PB117330BL	0.005	0.0	0.015	
PB117330BS	1.001	0.0	0.121	
K1560-01	0.016	0.0	0.016	
K1560-02	0.101	0.0	0.025	
PB117332BL	0.005	0.0	0.015	
PB117332BS	0.981	0.0	0.119	
K1560-07	0.016	0.0	0.016	
K1560-08	0.104	0.0	0.025	
CCV2	0.977	0.0	0.119	
CCB2	0.006	0.0	0.015	

N	14
Mean	0.371
SD	0.4776
CV%	128.69

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

2/27/2019 14:56


Test Phenolics-

Accepted 2/27/2019 13:48

Factor 9.35
Bias 0.014

Coeff. of det. 0.999908

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.019	0.0477	0.0500	
2	Phenol-2	0.025	0.1048	0.1000	
3	Phenol-2	0.042	0.2561	0.2500	
4	Phenol-2	0.067	0.4981	0.5000	
5	Phenol-2	0.120	0.9876	1.0000	
6	Phenol-2	0.229	2.0057	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 27 14:58:56 2019

Wed Feb 27 15:04:51 2019

A

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0477	mg/l	2/27/2019 13:47:03	
0.1PPM	A	Phenolics-	P	0.1048	mg/l	2/27/2019 13:47:04	
0.25PPM	A	Phenolics-	P	0.2561	mg/l	2/27/2019 13:47:05	
0.50PPM	A	Phenolics-	P	0.4981	mg/l	2/27/2019 13:47:06	
1.0PPM	A	Phenolics-	P	0.9876	mg/l	2/27/2019 13:47:07	
2.0PPM	A	Phenolics-	P	2.0057	mg/l	2/27/2019 13:47:08	
ICV1	S	Phenolics-	P	1.0016	mg/l	2/27/2019 14:58:56	
ICB1	S	Phenolics-	P	0.002	mg/l	2/27/2019 14:58:57	
CCV1	S	Phenolics-	P	0.9751	mg/l	2/27/2019 14:58:58	
CCB1	S	Phenolics-	P	0.0053	mg/l	2/27/2019 14:58:59	
PB117330BL	S	Phenolics-	P	0.0053	mg/l	2/27/2019 14:59:00	
PB117330BS	S	Phenolics-	P	1.0006	mg/l	2/27/2019 14:59:01	
K1560-01	S	Phenolics-	P	0.0161	mg/l	2/27/2019 14:59:02	
K1560-02	S	Phenolics-	P	0.1007	mg/l	2/27/2019 14:59:05	
PB117332BL	S	Phenolics-	P	0.0055	mg/l	2/27/2019 14:59:06	
PB117332BS	S	Phenolics-	P	0.9811	mg/l	2/27/2019 14:59:07	
K1560-07	S	Phenolics-	P	0.0157	mg/l	2/27/2019 15:04:46	
K1560-08	S	Phenolics-	P	0.1041	mg/l	2/27/2019 15:04:48	
CCV2	S	Phenolics-	P	0.9773	mg/l	2/27/2019 15:04:50	
CCB2	S	Phenolics-	P	0.0058	mg/l	2/27/2019 15:04:51	