

LB101055


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs. :

61453



A= 0.0000e+000

B= 1.6182e-004

C= 1.6532e-003

Rho= 0.9998921

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	63	0.000	63				
0.2	0.200	0.201	0.001	1231	0.0 %	1231				
2.5	2.500	2.419	-0.081	14937	0.0 %	14937				
5.0	5.000	5.078	0.078	31367	0.0 %	31367				
7.5	7.500	7.545	0.045	46612	0.0 %	46612				
10.0	10.000	9.946	-0.054	61453	0.0 %	61453				

LB101055 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Feb 2019 16:12:03

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	63	-	0.0000	7471B	PPB	27 Feb 2019 16:20:51	S	Std
0.2 - 1	50.2	1231	-	0.2000	7471B	PPB	27 Feb 2019 16:23:08	S	Std
2.5 - 1	52.5	14937	-	2.5000	7471B	PPB	27 Feb 2019 16:25:25	S	Std
5.0 - 1	55.0	31367	-	5.0000	7471B	PPB	27 Feb 2019 16:27:42	S	Std
7.5 - 1	57.5	46612	-	7.5000	7471B	PPB	27 Feb 2019 16:29:59	S	Std
10.0 - 1	510.0	61453	-	10.0000	7471B	PPB	27 Feb 2019 16:42:48	S	Std
ICV29 - 1	ICV29	23320	3.7754	-	7471B	PPB	27 Feb 2019 16:55:05	U	SMPL
ICB29 - 1	ICB29	-23	-0.0021	-	7471B	PPB	27 Feb 2019 16:57:20	U	SMPL
CCV39 - 1	CCV39	30395	4.9203	-	7471B	PPB	27 Feb 2019 16:59:36	U	SMPL
CCB39 - 1	CCB39	-71	-0.0098	-	7471B	PPB	27 Feb 2019 17:01:51	U	SMPL
CRA - 1	CRA	1140	0.1861	-	7471B	PPB	27 Feb 2019 17:04:08	U	SMPL
HighStd - 1	HighStd	61815	10.0048	-	7471B	PPB	27 Feb 2019 17:06:23	U	SMPL
ChkStd - 1	ChkStd	42311	6.8486	-	7471B	PPB	27 Feb 2019 17:08:38	U	SMPL
PB117443BL - 1	PBS	-28	-0.0029	-	7471B	PPB	27 Feb 2019 17:10:56	U	SMPL
PB117443BS - 1	LCSS	21901	3.5458	-	7471B	PPB	27 Feb 2019 17:13:14	U	SMPL
K1350-11 - 1	EO-01-022619-A	3649	0.5922	-	7471B	PPB	27 Feb 2019 17:15:30	U	SMPL
K1350-13 - 1	EO-01-022619-B	5061	0.8206	-	7471B	PPB	27 Feb 2019 17:17:47	U	SMPL
K1350-15 - 1	EO-01-022619-C	5100	0.8270	-	7471B	PPB	27 Feb 2019 17:20:04	U	SMPL
K1594-21 - 1	B-1(4-5)	1016	0.1661	-	7471B	PPB	27 Feb 2019 17:22:21	U	SMPL
K1594-22 - 1	B-1(8-9)	4963	0.8048	-	7471B	PPB	27 Feb 2019 17:24:37	U	SMPL
CCV40 - 1	CCV40	30601	4.9536	-	7471B	PPB	27 Feb 2019 17:26:52	U	SMPL
CCB40 - 1	CCB40	-79	-0.0111	-	7471B	PPB	27 Feb 2019 17:29:08	U	SMPL
K1594-24 - 1	B-2(8-9)	4168	0.6761	-	7471B	PPB	27 Feb 2019 17:31:24	U	SMPL
K1594-26 - 1	B-2B(21-22)	127	0.0222	-	7471B	PPB	27 Feb 2019 17:33:40	U	SMPL
K1594-28 - 1	B-3(8-9)	179	0.0306	-	7471B	PPB	27 Feb 2019 17:35:56	U	SMPL
K1594-29 - 1	B-4(4-5)	404	0.0670	-	7471B	PPB	27 Feb 2019 17:38:12	U	SMPL
K1594-31 - 1	B-4B(14-15)	171	0.0293	-	7471B	PPB	27 Feb 2019 17:40:28	U	SMPL
K1594-34 - 1	B-6(9-10)	525	0.0866	-	7471B	PPB	27 Feb 2019 17:42:44	U	SMPL
K1594-38 - 1	B-8(7.5-8.5)	695	0.1141	-	7471B	PPB	27 Feb 2019 17:45:01	U	SMPL
K1594-39 - 1	B-9(4-5)	4035	0.6546	-	7471B	PPB	27 Feb 2019 17:47:18	U	SMPL
K1594-40 - 1	B-9(7.5-8.5)	913	0.1494	-	7471B	PPB	27 Feb 2019 17:49:36	U	SMPL
K1620-02 - 1	SB-01	223	0.0377	-	7471B	PPB	27 Feb 2019 17:51:51	U	SMPL
CCV41 - 1	CCV41	29698	4.8075	-	7471B	PPB	27 Feb 2019 17:54:07	U	SMPL
CCB41 - 1	CCB41	-89	-0.0127	-	7471B	PPB	27 Feb 2019 17:56:23	U	SMPL
K1620-04 - 1	SB-02	448	0.0742	-	7471B	PPB	27 Feb 2019 17:58:40	U	SMPL
K1620-06 - 1	SB-03	478	0.0790	-	7471B	PPB	27 Feb 2019 18:00:55	U	SMPL
K1620-08 - 1	SB-04	2062	0.3353	-	7471B	PPB	27 Feb 2019 18:03:11	U	SMPL
K1620-10 - 1	SB-05	3048	0.4949	-	7471B	PPB	27 Feb 2019 18:05:27	U	SMPL
K1620-10DUP - 1	SB-05DUP	3037	0.4931	-	7471B	PPB	27 Feb 2019 18:07:43	U	SMPL
K1620-10MS - 1	SB-05MS	26555	4.2989	-	7471B	PPB	27 Feb 2019 18:09:59	U	SMPL
K1620-10MSD - 1	SB-05MSD	26447	4.2814	-	7471B	PPB	27 Feb 2019 18:12:16	U	SMPL
PB117444BL - 1	PBS	48	0.0094	-	7471B	PPB	27 Feb 2019 18:14:34	U	SMPL
PB117444BS - 1	LCSS	22174	3.5899	-	7471B	PPB	27 Feb 2019 18:16:52	U	SMPL
K1353-01 - 1	PL-02-022619-A	1137	0.1856	-	7471B	PPB	27 Feb 2019 18:19:08	U	SMPL
CCV42 - 1	CCV42	30330	4.9098	-	7471B	PPB	27 Feb 2019 18:21:24	U	SMPL
CCB42 - 1	CCB42	-77	-0.0108	-	7471B	PPB	27 Feb 2019 18:23:40	U	SMPL
K1353-03 - 1	PL-02-022619-B	2857	0.4640	-	7471B	PPB	27 Feb 2019 18:25:57	U	SMPL
K1353-05 - 1	PL-02-022619-C	1961	0.3190	-	7471B	PPB	27 Feb 2019 18:28:13	U	SMPL
K1353-07 - 1	PL-02-022619-D	3239	0.5258	-	7471B	PPB	27 Feb 2019 18:30:29	U	SMPL
K1627-01 - 1	SB-01-(0-6)	7597	1.2310	-	7471B	PPB	27 Feb 2019 18:32:45	U	SMPL
K1627-02 - 1	SB-01-(6-2)	57604	9.3234	-	7471B	PPB	27 Feb 2019 18:35:00	U	SMPL
K1627-03 - 1	SB-02-(2-5)	3763	0.6106	-	7471B	PPB	27 Feb 2019 18:37:17	U	SMPL
K1627-04 - 1	SB-02-(5-10)	2423	0.3938	-	7471B	PPB	27 Feb 2019 18:39:35	U	SMPL
K1627-05 - 1	SB-03-(0-2)	12550	2.0326	-	7471B	PPB	27 Feb 2019 18:41:52	U	SMPL
K1627-06 - 1	SB-03-(2-5)	11908	1.9287	-	7471B	PPB	27 Feb 2019 18:44:09	U	SMPL
K1627-07 - 1	SB-04-(0-2)	11477	1.8589	-	7471B	PPB	27 Feb 2019 18:46:26	U	SMPL
CCV43 - 1	CCV43	29633	4.7970	-	7471B	PPB	27 Feb 2019 18:48:45	U	SMPL
CCB43 - 1	CCB43	-87	-0.0124	-	7471B	PPB	27 Feb 2019 18:51:01	U	SMPL
K1627-08 - 1	SB-04-(2-5)	6311	1.0229	-	7471B	PPB	27 Feb 2019 18:53:18	U	SMPL
K1633-03 - 1	SR-11-S	1029	0.1682	-	7471B	PPB	27 Feb 2019 18:55:34	U	SMPL
K1633-07 - 1	SR-11-D	1280	0.2088	-	7471B	PPB	27 Feb 2019 18:57:50	U	SMPL
K1633-07DUP - 1	SR-11-DDUP	1290	0.2104	-	7471B	PPB	27 Feb 2019 19:00:05	U	SMPL
K1633-07MS - 1	SR-11-DMS	25601	4.1445	-	7471B	PPB	27 Feb 2019 19:06:55	U	SMPL
K1633-07MSD - 1	SR-11-DMSD	24952	4.0395	-	7471B	PPB	27 Feb 2019 19:09:11	U	SMPL
K1620-10LX5 - 1		512	0.0845	-	7471B	PPB	27 Feb 2019 19:11:29	U	SMPL
K1620-10A - 1		26733	4.3277	-	7471B	PPB	27 Feb 2019 19:13:47	U	SMPL
K1633-07LX5 - 1		182	0.0311	-	7471B	PPB	27 Feb 2019 19:16:04	U	SMPL
K1633-07A - 1		25785	4.1743	-	7471B	PPB	27 Feb 2019 19:18:21	U	SMPL
CCV44 - 1	CCV44	30267	4.8996	-	7471B	PPB	27 Feb 2019 19:20:37	U	SMPL
CCB44 - 1	CCB44	28	0.0062	-	7471B	PPB	27 Feb 2019 19:22:55	U	SMPL