



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101064

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	02/27/2019	10:05
2	K1631-14		1	7	>100.0	>212.0	>212.0	02/27/2019	10:40
3	K1631-14DUP		1	7	>100.0	>212.0	>212.0	02/27/2019	11:05
4	K1632-02		1	8	>100.0	>212.0	>212.0	02/27/2019	11:40
5	K1632-03		1	8	>100.0	>212.0	>212.0	02/27/2019	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB101064

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP22719

WorkList ID : 122770

Date : 02-27-2019 09:26:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2019	Water	K1631-14 C	Flash Point	Cool 4 deg C	PSEG01	K31	Z31W	02/26/2019	1010A
03/01/2019	Water	K1632-02 B	Flash Point	Cool 4 deg C	PSEG01	K31	30236	02/26/2019	1010A
03/01/2019	Water	K1632-03 D	Flash Point	Cool 4 deg C	PSEG01	K31	30238	02/26/2019	1010A

Date/Time 02/27/19 9:30

Received by: KS

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

02/27/19 12:15

CP

KS