



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 02/28/2019

Run Number: LB101067

BalanceID: WC SC-3

OvenID: WC Oven-2

FilterID: 14466061

ThermometerID: WET-OVEN#2

TEMP1 IN: 103 02/27/2019 16:00 TEMP OUT: 104 02/28/2019 07:00
TEMP2 IN: 104 02/28/2019 07:35 TEMP OUT: 104 02/28/2019 08:45
TEMP3 IN: 105 02/28/2019 09:30 TEMP OUT: 105 02/28/2019 11:10
TEMP4 IN: 105 02/28/2019 12:00 TEMP OUT: 105 02/28/2019 13:40

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB101067BL	LB101067BL	1.4033	1.4033	100	1.4033	1.4033	1.4033	0.0000	0
2	LB101067BS	LB101067BS	1.4056	1.4056	100	1.4558	1.4558	1.4558	0.0502	502
3	K1629-01	001M-WILLETS-PT-BLVD (FEB)	1.4110	1.4110	200	1.4198	1.4198	1.4198	0.0088	44
4	K1642-01	COMP	1.3863	1.3863	50	1.4264	1.4264	1.4264	0.0401	802
5	K1642-01DUP	COMPDUP	1.4084	1.4084	50	1.4492	1.4492	1.4492	0.0408	816

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS022819

WorkList ID : 122807

Date : 02-28-2019 08:16:14

23101067

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2019	Water	K1629-01	✓ TSS	Cool 4 deg C	TULL01	P13	001M-WILLETS-PT-BLVD(FEI	02/26/2019	SM2540 D
03/06/2019	Water	K1642-01	✓ TSS	Cool 4 deg C	ARAM01	P33	COMP	02/27/2019	SM2540 D

Date/Time 02/28/19, 08:20
 Received by: AS
 Relinquished by: [Signature]

Date/Time 02/28/19, 17:30
 Received by: [Signature]
 Relinquished by: AS