



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101087

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/28/2019	16:05
2	CAL2	1	Water	NA	NA	20.2	7.01	02/28/2019	16:06
3	CAL3	1	Water	NA	NA	20.2	10.04	02/28/2019	16:07
4	ICV	1	Water	NA	NA	20.1	7.00	02/28/2019	16:08
5	CCV1	1	Water	NA	NA	21.2	2.01	02/28/2019	16:10
6	K1655-01	1	Solid	20.02	20	21.6	8.44	02/28/2019	16:11
7	K1655-02	1	Solid	20.03	20	21.7	8.30	02/28/2019	16:12
8	K1655-02DUP	1	Solid	20.04	20	21.6	8.31	02/28/2019	16:13
9	K1655-10	1	Solid	20.02	20	21.6	7.02	02/28/2019	16:15
10	CCV2	1	Water	NA	NA	20.2	12.01	02/28/2019	16:16

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1655

WorkList ID : 122860

Date : 02-28-2019 15:40:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2019	Solid	K1655-01	pH	Cool 4 deg C	WEST10	P22	A0C12-1-1.0-1.5	02/27/2019	9045D
02/27/2019	Solid	K1655-02	pH	Cool 4 deg C	WEST10	P22	DUP-2	02/27/2019	9045D
02/27/2019	Solid	K1655-10	pH	Cool 4 deg C	WEST10	P22	A0C12-2-1.5-2.0	02/27/2019	9045D

Date/Time 02-28-19 16:00

Received by: Si

Relinquished by: CC

Date/Time 02-28-19 17:00

Received by: SP

Relinquished by: 5