



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB101088

Analyst By : jignesh
Supervisor Review By : apatel
Slope : 98.5
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/28/2019	16:05
2	CAL2	1	Water	NA	NA	20.2	7.01	02/28/2019	16:06
3	CAL3	1	Water	NA	NA	20.2	10.04	02/28/2019	16:07
4	ICV	1	Water	NA	NA	20.1	7.00	02/28/2019	16:08
5	CCV1	1	Water	NA	NA	20.2	2.01	02/28/2019	16:20
6	K1655-04	1	Water	NA	NA	18.4	6.80	02/28/2019	16:21
7	K1674-02	1	Water	NA	NA	14.0	6.69	02/28/2019	16:22
8	K1674-03	1	Water	NA	NA	14.6	5.99	02/28/2019	16:24
9	K1674-04	1	Water	NA	NA	17.6	6.14	02/28/2019	16:25
10	K1674-05	1	Water	NA	NA	18.2	6.22	02/28/2019	16:26
11	K1674-05DUP	1	Water	NA	NA	18.3	6.23	02/28/2019	16:27
12	CCV2	1	Water	NA	NA	20.2	12.01	02/28/2019	16:30

LB 101088

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1673

WorkList ID : 122859

Date : 02-28-2019 15:38:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2019	Water	K1655-04-11	pH	Cool 4 deg C	WEST10	P22	FB-022719	02/27/2019	9041A
02/28/2019	Water	K1674-02	pH	Cool 4 deg C	PSEG01	K31	A	02/28/2019	9040C
02/28/2019	Water	K1674-03	pH	Cool 4 deg C	PSEG01	K31	B	02/28/2019	9040C
02/28/2019	Water	K1674-04	pH	Cool 4 deg C	PSEG01	K31	C	02/28/2019	9040C
02/28/2019	Water	K1674-05	pH	Cool 4 deg C	PSEG01	K31	D	02/28/2019	9040C

Date/Time 02-28-19 16:00

Received by: JD

Relinquished by: CA

Date/Time 02-28-19 17:10

Received by: JP

Relinquished by: CA