

LB101096


ISM02.4

INSTRUMENT ID : CUI

Linear

 μ Abs. :

47622



A= 0.0000e+000

B= 2.0930e-004

C= -1.1100e-002

Rho= 0.9999437

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.006	-0.006	26	0.000	26				
0.2	0.200	0.196	-0.004	990	0.0 %	990				
2.5	2.500	2.512	0.012	12054	0.0 %	12054				
5.0	5.000	4.965	-0.035	23773	0.0 %	23773				
7.5	7.500	7.577	0.077	36252	0.0 %	36252				
10	10.000	9.956	-0.044	47622	0.0 %	47622				

%D
-2
0
-1
1
0

LB101096
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	26	-		0 ISM02.4	PPB	3/1/2019 8:53	Std
	0.2 S0.2	990	-		0.2 ISM02.4	PPB	3/1/2019 8:55	Std
	2.5 S2.5	12054	-		2.5 ISM02.4	PPB	3/1/2019 8:57	Std
	5 S5.0	23773	-		5 ISM02.4	PPB	3/1/2019 9:00	Std
	7.5 S7.5	36252	-		7.5 ISM02.4	PPB	3/1/2019 9:02	Std
	10 S10.0	47622	-		10 ISM02.4	PPB	3/1/2019 9:04	Std
ICV	ICV	18763	3.9161 -		ISM02.4	PPB	3/1/2019 9:09	SMPL
ICB	ICB	26	-0.0057 -		ISM02.4	PPB	3/1/2019 9:11	SMPL
CCV058	CCV058	23958	5.0034 -		ISM02.4	PPB	3/1/2019 9:13	SMPL
CCB058	CCB058	-16	-0.0144 -		ISM02.4	PPB	3/1/2019 9:15	SMPL
PB117489BL	PBW489	170	0.0245 -		ISM02.4	PPB	3/1/2019 9:18	SMPL
K1554-07	MX2G09	30568	6.3869 -		ISM02.4	PPB	3/1/2019 9:22	SMPL
K1554-08	MX2G09D	30274	6.3254 -		ISM02.4	PPB	3/1/2019 9:24	SMPL
K1554-09	MX2G09S	34750	7.2622 -		ISM02.4	PPB	3/1/2019 9:27	SMPL
CCV059	CCV059	23856	4.9821 -		ISM02.4	PPB	3/1/2019 9:29	SMPL
CCB059	CCB059	15	-0.008 -		ISM02.4	PPB	3/1/2019 9:31	SMPL