

LB101114


7471B

INSTRUMENT 10: CV1

Linear

 μ Abs.:

57520



A= 0.0000e+000

B= 1.7298e-004

C= -5.2594e-002

Rho= 0.9998440

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.048	-0.048	26	0.000	26				
0.2	0.200	0.176	-0.024	1324	0.0 %	1324				
2.5	2.500	2.523	0.023	14887	0.0 %	14887				
5.0	5.000	5.075	0.075	29645	0.0 %	29645				
7.5	7.500	7.577	0.077	44104	0.0 %	44104				
10.0	10.000	9.897	-0.103	57520	0.0 %	57520				

LB101114 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Mar 2019 14:11:59

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	26	-	0.0000	7471B	PPB	01 Mar 2019 14:14:05	S	Std
0.2 - 1	50-2	1324	-	0.2000	7471B	PPB	01 Mar 2019 14:16:22	S	Std
2.5 - 1	52-5	14887	-	2.5000	7471B	PPB	01 Mar 2019 14:18:40	S	Std
5.0 - 1	55-0	29645	-	5.0000	7471B	PPB	01 Mar 2019 14:20:57	S	Std
7.5 - 1	57-5	44104	-	7.5000	7471B	PPB	01 Mar 2019 14:23:13	S	Std
10.0 - 1	510-0	57520	-	10.0000	7471B	PPB	01 Mar 2019 14:25:30	S	Std
ICV36 - 1	ICV36	23022	3.9297	-	7471B	PPB	01 Mar 2019 14:31:13	U	SMPL
ICB36 - 1	ICB36	-52	-0.0616	-	7471B	PPB	01 Mar 2019 14:33:29	U	SMPL
CCV66 - 1	CCV66	29635	5.0737	-	7471B	PPB	01 Mar 2019 14:35:46	U	SMPL
CCB66 - 1	CCB66	-83	-0.0670	-	7471B	PPB	01 Mar 2019 14:38:01	U	SMPL
CRA - 1	CRA	1284	0.1695	-	7471B	PPB	01 Mar 2019 14:40:17	U	SMPL
HighStd - 1	HighStd	57987	9.9780	-	7471B	PPB	01 Mar 2019 14:42:32	U	SMPL
ChkStd - 1	ChkStd	40526	6.9576	-	7471B	PPB	01 Mar 2019 14:44:48	U	SMPL
PB117520BL - 1	PBS	110	-0.0336	-	7471B	PPB	01 Mar 2019 14:47:06	U	SMPL
PB117520BS - 1	LCSS	21287	3.6296	-	7471B	PPB	01 Mar 2019 14:49:24	U	SMPL
K1669-01 - 1	FK-GF-02-031-A	5758	0.9434	-	7471B	PPB	01 Mar 2019 14:51:40	U	SMPL
K1669-01DUP - 1	FK-GF-02-031-ADUP	5856	0.9604	-	7471B	PPB	01 Mar 2019 14:53:57	U	SMPL
K1669-01MS - 1	FK-GF-02-031-AMS	25815	4.4129	-	7471B	PPB	01 Mar 2019 14:56:14	U	SMPL
K1669-01MSD - 1	FK-GF-02-031-AMSD	25559	4.3686	-	7471B	PPB	01 Mar 2019 14:58:31	U	SMPL
K1669-03 - 1	FK-GF-02-031-B	3682	0.5843	-	7471B	PPB	01 Mar 2019 15:00:47	U	SMPL
CCV67 - 1	CCV67	29521	5.0539	-	7471B	PPB	01 Mar 2019 15:03:04	U	SMPL
CCB67 - 1	CCB67	143	-0.0279	-	7471B	PPB	01 Mar 2019 15:05:19	U	SMPL
K1670-01 - 1	FK-GF-02-032-A	3676	0.5833	-	7471B	PPB	01 Mar 2019 15:07:35	U	SMPL
K1670-03 - 1	FK-GF-02-032-B	3262	0.5117	-	7471B	PPB	01 Mar 2019 15:09:51	U	SMPL
K1671-01 - 1	FK-BPM-04-002-ABCD	1954	0.2854	-	7471B	PPB	01 Mar 2019 15:12:07	U	SMPL
K1672-01 - 1	FK-GF-02-033-A	4486	0.7234	-	7471B	PPB	01 Mar 2019 15:14:23	U	SMPL
K1672-03 - 1	FK-GF-02-033-B	3907	0.6232	-	7471B	PPB	01 Mar 2019 15:16:39	U	SMPL
K1675-01 - 1	Z-3-11-A	2279	0.3416	-	7471B	PPB	01 Mar 2019 15:18:55	U	SMPL
K1675-05 - 1	Z-3-11-B	769	0.0804	-	7471B	PPB	01 Mar 2019 15:21:11	U	SMPL
K1675-09 - 1	Z-3-1-A	2555	0.3894	-	7471B	PPB	01 Mar 2019 15:23:29	U	SMPL
K1675-11 - 1	Z-3-1-B	24760	4.2304	-	7471B	PPB	01 Mar 2019 15:25:46	U	SMPL
K1675-13 - 1	Z-3-2-A	13122	2.2172	-	7471B	PPB	01 Mar 2019 15:28:02	U	SMPL
CCV68 - 1	CCV68	29616	5.0704	-	7471B	PPB	01 Mar 2019 15:30:18	U	SMPL
CCB68 - 1	CCB68	84	-0.0381	-	7471B	PPB	01 Mar 2019 15:32:33	U	SMPL
K1675-15 - 1	Z-3-2-B	10743	1.8057	-	7471B	PPB	01 Mar 2019 15:34:49	U	SMPL
K1675-17 - 1	Z-4-A-B	127650	22.0283	-	7471B	PPB	01 Mar 2019 15:37:06	U	SMPL
K1675-23 - 1	Z-3-12-A	332	0.0048	-	7471B	PPB	01 Mar 2019 15:39:21	U	SMPL
K1675-27 - 1	Z-3-12-B	5505	0.8997	-	7471B	PPB	01 Mar 2019 15:41:40	U	SMPL
K1675-29 - 1	Z-3-13-A	5645	0.9239	-	7471B	PPB	01 Mar 2019 15:43:56	U	SMPL
K1675-33 - 1	Z-3-13-B	1864	0.2698	-	7471B	PPB	01 Mar 2019 15:46:12	U	SMPL
K1675-37 - 1	Z-3-14-A	13491	2.2811	-	7471B	PPB	01 Mar 2019 15:48:29	U	SMPL
K1675-41 - 1	Z-3-14-B	699	0.0683	-	7471B	PPB	01 Mar 2019 15:50:46	U	SMPL
PB117521BL - 1	PBS	106	-0.0343	-	7471B	PPB	01 Mar 2019 15:53:03	U	SMPL
PB117521BS - 1	LCSS	21635	3.6898	-	7471B	PPB	01 Mar 2019 15:55:19	U	SMPL
CCV69 - 1	CCV69	28891	4.9450	-	7471B	PPB	01 Mar 2019 16:03:07	U	SMPL
CCB69 - 1	CCB69	-58	-0.0626	-	7471B	PPB	01 Mar 2019 16:05:22	U	SMPL
K1655-01 - 1	A0C12-1-1.0-1.5	1391	0.1880	-	7471B	PPB	01 Mar 2019 16:07:39	U	SMPL
K1655-02 - 1	DUP-2	1206	0.1560	-	7471B	PPB	01 Mar 2019 16:09:55	U	SMPL
K1655-10 - 1	A0C12-2-1.5-2.0	2412	0.3646	-	7471B	PPB	01 Mar 2019 16:12:10	U	SMPL
K1655-10DUP - 1	A0C12-2-1.5-2.0DUP	2380	0.3591	-	7471B	PPB	01 Mar 2019 16:14:26	U	SMPL
K1655-11 - 1	K1655-10MS	24158	4.1263	-	7471B	PPB	01 Mar 2019 16:16:42	U	SMPL
K1655-12 - 1	K1655-10MSD	23954	4.0910	-	7471B	PPB	01 Mar 2019 16:18:58	U	SMPL
K1665-01 - 1	200051	2247	0.3361	-	7471B	PPB	01 Mar 2019 16:21:16	U	SMPL
K1665-03 - 1	200029	1476	0.2027	-	7471B	PPB	01 Mar 2019 16:23:34	U	SMPL
K1665-05 - 1	200053	2202	0.3283	-	7471B	PPB	01 Mar 2019 16:25:51	U	SMPL
K1665-07 - 1	200031	2773	0.4271	-	7471B	PPB	01 Mar 2019 16:28:06	U	SMPL
CCV70 - 1	CCV70	29928	5.1243	-	7471B	PPB	01 Mar 2019 16:30:22	U	SMPL
CCB70 - 1	CCB70	-72	-0.0650	-	7471B	PPB	01 Mar 2019 16:32:38	U	SMPL
K1669-01LX5 - 1		1131	0.1430	-	7471B	PPB	01 Mar 2019 16:34:55	U	SMPL
K1669-01A - 1		25904	4.4283	-	7471B	PPB	01 Mar 2019 16:37:11	U	SMPL
K1655-10LX5 - 1		370	0.0114	-	7471B	PPB	01 Mar 2019 16:39:26	U	SMPL
K1655-10A - 1		24205	4.1344	-	7471B	PPB	01 Mar 2019 16:41:44	U	SMPL
K1675-17DLX5 - 1		25386	4.3387	-	7471B	PPB	01 Mar 2019 16:43:59	U	SMPL
CCV71 - 1	CCV71	29681	5.0816	-	7471B	PPB	01 Mar 2019 16:46:17	U	SMPL
CCB71 - 1	CCB71	-30	-0.0578	-	7471B	PPB	01 Mar 2019 16:48:35	U	SMPL