



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB101121

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/01/2019	17:35
2	CAL2	1	Water	NA	NA	20.4	7.01	03/01/2019	17:36
3	CAL3	1	Water	NA	NA	20.4	10.02	03/01/2019	17:37
4	ICV	1	Water	NA	NA	20.4	7.00	03/01/2019	17:38
5	CCV1	1	Water	NA	NA	20.5	2.02	03/01/2019	17:40
6	K1700-02	1	Water	NA	NA	11.5	6.24	03/01/2019	17:42
7	K1715-14	1	Water	NA	NA	13.3	6.66	03/01/2019	17:43
8	K1715-14DUP	1	Water	NA	NA	13.4	6.67	03/01/2019	17:44
9	CCV2	1	Water	NA	NA	20.5	12.02	03/01/2019	17:45

2B 101121

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K1715 WorkList ID : 122928 Date : 03-01-2019 17:25:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/01/2019	Water - A	K1700-02	Corrosivity	Cool 4 deg C	PSEG01	K31	NB-21841-21842	03/01/2019	9040C
03/01/2019	Water - B	K1715-14	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-4-B	03/01/2019	9040C

Date/Time 03-01-19 17:30
 Received by: Jy
 Relinquished by: CO

Date/Time 03-01-19 18:30
 Received by: Jy
 Relinquished by: CO