



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB101128
Analysis Date: 03/04/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 03/04/2019
Extraction IN Time: 07:30
Extraction OUT Time: 08:10

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB101128BL	LB101128BL	WATER	2	1000	100	3.1477	3.1477	3.02	3.1477	3.1477	0.0000	0
2	LB101128BS	LB101128BS	WATER	2	1000	100	3.5698	3.5698	3.01	3.5888	3.5888	0.0190	19
3	K1700-01	NB-21841-21842	WATER	2	1000	100	3.0087	3.0087	18.04	3.4286	3.4286	0.4199	419.9
4	K1715-13	TP-4-B	WATER	2	1000	100	2.9466	2.9466	3.03	2.9470	2.9470	0.0004	0.4
5	K1716-01	STORM-DRAIN-1	WATER	2	1000	100	2.9876	2.9876	3.04	2.9885	2.9885	0.0009	0.9
6	K1716-01MS	STORM-DRAIN-1	WATER	2	1000	100	3.9657	3.9657	3.03	3.9865	3.9865	0.0208	20.8
7	K1716-02	STORM-DRAIN-2	WATER	2	1000	100	3.0560	3.0560	3.02	3.0577	3.0577	0.0017	1.7
8	K1716-02DUP	STORM-DRAIN-2DUP	WATER	2	1000	100	3.1811	3.1811	3.01	3.1830	3.1830	0.0019	1.9



QC Batch# LB101128 Test: TPH Analysis Date: 03/04/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2494
pH Paper 0-14	M4240
Sodium Sulfate	EP1930
1:1 HCL	WP69312
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP71442
LCSWD	NA	NA
MS/MSD	10.00 ML	WP71443

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:36</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:35</u>		
Bal Check Time:	<u>07:35</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:10</u>
		Out Time:	<u>09:35</u>		

After Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:16</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>10:35</u>		
Bal Check Time:	<u>12:15</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:40</u>
		Out Time:	<u>11:15</u>		

LB 101128

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph k1716

WorkList ID : 122947

Date : 03-04-2019 7:20:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/08/2019	Water	K1700-01-H	TPH	Conc H2SO4 tr	PSEG01	K31	NB-21841-21842	03/01/2019	1664A
03/06/2019	Water	K1715-13 -H	TPH	Conc H2SO4 tr	PSEG01	K31	TP-4-B	03/01/2019	1664A
03/08/2019	Water	K1716-01 E, F	TPH	Conc H2SO4 tr	KWAR01	P51	STORM-DRAIN-1	03/01/2019	1664A
03/08/2019	Water	K1716-02	TPH	Conc H2SO4 tr	KWAR01	P51	STORM-DRAIN-2	03/01/2019	1664A

E, F,

Date/Time 03-04-19 7:25

Received by: Jn

Relinquished by: CB

Date/Time 03-04-19 9:00

Received by: CB

Relinquished by: Jn