

LB101135


7470A

INSTRUMENT ID: CU1

Linear

 μ Abs.:

42792



A= 0.0000e+000

B= 2.3573e-004

C= -1.0653e-002

Rho= 0.9997653

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-5	0.000	-5				
0.2	0.200	0.224	0.024	995	0.0 %	995				
2.5	2.500	2.488	-0.012	10601	0.0 %	10601				
5.0	5.000	5.081	0.081	21600	0.0 %	21600				
7.5	7.500	7.342	-0.158	31190	0.0 %	31190				
10.0	10.000	10.077	0.077	42792	0.0 %	42792				

LB101135 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Mar 2019 11:23:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-5	-	0.0000	7470A	PPB	04 Mar 2019 11:27:23	S	Std
0.2 - 1	50-2	995	-	0.2000	7470A	PPB	04 Mar 2019 11:29:31	S	Std
2.5 - 1	52-5	10601	-	2.5000	7470A	PPB	04 Mar 2019 11:31:39	S	Std
5.0 - 1	55-0	21600	-	5.0000	7470A	PPB	04 Mar 2019 11:33:46	S	Std
7.5 - 1	57-5	31190	-	7.5000	7470A	PPB	04 Mar 2019 11:35:54	S	Std
10.0 - 1	510-0	42792	-	10.0000	7470A	PPB	04 Mar 2019 11:38:00	S	Std
ICV37 - 1	ICV37	17000	3.9968	-	7470A	PPB	04 Mar 2019 11:41:38	U	SMPL
ICB37 - 1	ICB37	31	-0.0033	-	7470A	PPB	04 Mar 2019 11:43:45	U	SMPL
CCV72 - 1	CCV72	21637	5.0898	-	7470A	PPB	04 Mar 2019 11:45:50	U	SMPL
CCB72 - 1	CCB72	-41	-0.0203	-	7470A	PPB	04 Mar 2019 11:47:56	U	SMPL
CRA - 1	CRA	955	0.2145	-	7470A	PPB	04 Mar 2019 11:50:02	U	SMPL
HighStd - 1	HighStd	43419	10.2245	-	7470A	PPB	04 Mar 2019 11:52:08	U	SMPL
ChkStd - 1	ChkStd	28965	6.8173	-	7470A	PPB	04 Mar 2019 11:54:14	U	SMPL
PB117579BL - 1	PBW	20	-0.0059	-	7470A	PPB	04 Mar 2019 11:56:22	U	SMPL
PB117579BS - 1	LCSW	17674	4.1556	-	7470A	PPB	04 Mar 2019 11:58:30	U	SMPL
K1665-02 - 1	200051	-3	-0.0114	-	7470A	PPB	04 Mar 2019 12:00:38	U	SMPL
K1665-02DUP - 1	200051DUP	17	-0.0066	-	7470A	PPB	04 Mar 2019 12:02:45	U	SMPL
K1665-02MS - 1	200051MS	17801	4.1856	-	7470A	PPB	04 Mar 2019 12:04:53	U	SMPL
K1665-02MSD - 1	200051MSD	17842	4.1952	-	7470A	PPB	04 Mar 2019 12:07:00	U	SMPL
K1665-04 - 1	200029	71	0.0061	-	7470A	PPB	04 Mar 2019 12:09:06	U	SMPL
CCV73 - 1	CCV73	21562	5.0722	-	7470A	PPB	04 Mar 2019 12:11:13	U	SMPL
CCB73 - 1	CCB73	-29	-0.0175	-	7470A	PPB	04 Mar 2019 12:13:19	U	SMPL
K1665-06 - 1	200053	71	0.0061	-	7470A	PPB	04 Mar 2019 12:15:26	U	SMPL
K1665-08 - 1	200031	51	0.0014	-	7470A	PPB	04 Mar 2019 12:17:32	U	SMPL
K1675-04 - 1	Z-3-11-A	55	0.0023	-	7470A	PPB	04 Mar 2019 12:19:38	U	SMPL
K1675-08 - 1	Z-3-11-B	58	0.0030	-	7470A	PPB	04 Mar 2019 12:21:45	U	SMPL
K1675-10 - 1	Z-3-1-A	163	0.0278	-	7470A	PPB	04 Mar 2019 12:23:54	U	SMPL
K1675-12 - 1	Z-3-1-B	130	0.0200	-	7470A	PPB	04 Mar 2019 12:26:01	U	SMPL
K1675-14 - 1	Z-3-2-A	147	0.0240	-	7470A	PPB	04 Mar 2019 12:28:08	U	SMPL
K1675-16 - 1	Z-3-2-B	149	0.0245	-	7470A	PPB	04 Mar 2019 12:30:16	U	SMPL
K1675-18 - 1	Z-4-A-B	148	0.0242	-	7470A	PPB	04 Mar 2019 12:32:24	U	SMPL
K1675-26 - 1	Z-3-12-A	94	0.0115	-	7470A	PPB	04 Mar 2019 12:34:31	U	SMPL
CCV74 - 1	CCV74	21698	5.1042	-	7470A	PPB	04 Mar 2019 12:36:37	U	SMPL
CCB74 - 1	CCB74	-65	-0.0260	-	7470A	PPB	04 Mar 2019 12:38:43	U	SMPL
K1675-28 - 1	Z-3-12-B	104	0.0139	-	7470A	PPB	04 Mar 2019 12:40:50	U	SMPL
K1675-32 - 1	Z-3-13-A	123	0.0183	-	7470A	PPB	04 Mar 2019 12:42:56	U	SMPL
K1675-36 - 1	Z-3-13-B	124	0.0186	-	7470A	PPB	04 Mar 2019 12:45:03	U	SMPL
K1675-40 - 1	Z-3-14-A	161	0.0273	-	7470A	PPB	04 Mar 2019 12:47:09	U	SMPL
K1675-44 - 1	Z-3-14-B	107	0.0146	-	7470A	PPB	04 Mar 2019 12:49:16	U	SMPL
PB117580BL - 1	PBW	104	0.0139	-	7470A	PPB	04 Mar 2019 12:51:24	U	SMPL
PB117580BS - 1	LCSW	17753	4.1743	-	7470A	PPB	04 Mar 2019 12:53:31	U	SMPL
K1666-01 - 1	PM-A	69	0.0056	-	7470A	PPB	04 Mar 2019 12:55:39	U	SMPL
K1666-02 - 1	PM-B	129	0.0198	-	7470A	PPB	04 Mar 2019 12:57:47	U	SMPL
K1666-03 - 1	PM-C	175	0.0306	-	7470A	PPB	04 Mar 2019 12:59:53	U	SMPL
CCV75 - 1	CCV75	21915	5.1554	-	7470A	PPB	04 Mar 2019 13:02:00	U	SMPL
CCB75 - 1	CCB75	-11	-0.0132	-	7470A	PPB	04 Mar 2019 13:04:07	U	SMPL
K1666-04 - 1	PM-D	121	0.0179	-	7470A	PPB	04 Mar 2019 13:06:14	U	SMPL
K1666-05 - 1	PM-E	131	0.0202	-	7470A	PPB	04 Mar 2019 13:08:20	U	SMPL
K1666-06 - 1	PM-F	126	0.0190	-	7470A	PPB	04 Mar 2019 13:10:26	U	SMPL
K1666-07 - 1	PM-G	136	0.0214	-	7470A	PPB	04 Mar 2019 13:12:33	U	SMPL
K1666-08 - 1	PM-H	111	0.0155	-	7470A	PPB	04 Mar 2019 13:14:40	U	SMPL
K1666-08DUP - 1	PM-HDUP	118	0.0172	-	7470A	PPB	04 Mar 2019 13:16:47	U	SMPL
K1666-08MS - 1	PM-HMS	18488	4.3475	-	7470A	PPB	04 Mar 2019 13:18:54	U	SMPL
K1666-08MSD - 1	PM-HMSD	18441	4.3364	-	7470A	PPB	04 Mar 2019 13:21:02	U	SMPL
PB117586BL - 1	PBW	44	-0.0003	-	7470A	PPB	04 Mar 2019 13:23:09	U	SMPL
PB117586BS - 1	LCSW	16164	3.7997	-	7470A	PPB	04 Mar 2019 13:25:16	U	SMPL
CCV76 - 1	CCV76	21455	5.0469	-	7470A	PPB	04 Mar 2019 13:27:23	U	SMPL
CCB76 - 1	CCB76	41	-0.0010	-	7470A	PPB	04 Mar 2019 13:29:31	U	SMPL
K1655-04 - 1	FB-022719	82	0.0087	-	7470A	PPB	04 Mar 2019 13:31:38	U	SMPL
K1700-01 - 1	NB-21841-21842	3199	0.7434	-	7470A	PPB	04 Mar 2019 13:33:44	U	SMPL
K1715-13 - 1	TP-4-B	825	0.1838	-	7470A	PPB	04 Mar 2019 13:35:51	U	SMPL
K1715-13DUP - 1	TP-4-BDUP	804	0.1789	-	7470A	PPB	04 Mar 2019 13:37:57	U	SMPL
K1715-13MS - 1	TP-4-BMS	19725	4.6391	-	7470A	PPB	04 Mar 2019 13:40:04	U	SMPL
K1715-13MSD - 1	TP-4-BMSD	19884	4.6766	-	7470A	PPB	04 Mar 2019 13:42:11	U	SMPL
PB117579TB - 1	PB117499TB	87	0.0099	-	7470A	PPB	04 Mar 2019 13:44:18	U	SMPL
PB117580TB - 1	PB117529TB	120	0.0176	-	7470A	PPB	04 Mar 2019 13:46:26	U	SMPL
K1665-02LX5 - 1		7	-0.0090	-	7470A	PPB	04 Mar 2019 13:48:34	U	SMPL
K1665-02A - 1		17669	4.1545	-	7470A	PPB	04 Mar 2019 13:50:41	U	SMPL
CCV77 - 1	CCV77	22090	5.1966	-	7470A	PPB	04 Mar 2019 13:52:47	U	SMPL
CCB77 - 1	CCB77	44	-0.0003	-	7470A	PPB	04 Mar 2019 13:54:54	U	SMPL
K1666-08LX5 - 1		17	-0.0066	-	7470A	PPB	04 Mar 2019 13:57:02	U	SMPL
K1666-08A - 1		19072	4.4852	-	7470A	PPB	04 Mar 2019 13:59:08	U	SMPL
K1715-13LX5 - 1		112	0.0157	-	7470A	PPB	04 Mar 2019 14:01:15	U	SMPL
K1715-13A - 1		19752	4.6455	-	7470A	PPB	04 Mar 2019 14:03:22	U	SMPL
CCV78 - 1	CCV78	21515	5.0611	-	7470A	PPB	04 Mar 2019 14:05:29	U	SMPL
CCB78 - 1	CCB78	-33	-0.0184	-	7470A	PPB	04 Mar 2019 14:07:36	U	SMPL