



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101142

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	03/04/2019	10:45
2	K1700-02		1	7	>100.0	>212.0	>212.0	03/04/2019	11:10
3	K1700-02DUP		1	7	>100.0	>212.0	>212.0	03/04/2019	11:40
4	K1715-14		1	8	>100.0	>212.0	>212.0	03/04/2019	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 101142

WorkList Name : FP3419

WorkList ID : 122972

Date : 03-04-2019 10:12:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/08/2019	Water	K1700-02 C	Flash Point	Cool 4 deg C	PSEG01	K31	NB-21841-21842	03/01/2019	1010A
03/06/2019	Water	K1715-14 B	Flash Point	Cool 4 deg C	PSEG01	K31	TP-4-B	03/01/2019	1010A

Date/Time 03/04/19 10:15
 Received by: KS
 Relinquished by: CP

Date/Time 03/04/19 12:15
 Received by: CP
 Relinquished by: KS