


LB101144

INSTRUMENT ID : CU1

245.1

 μ Abs. :

40824



Linear

A= 0.0000e+000

B= 2.4382e-004

C= -1.0481e-002

Rho= 0.9999454

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	35	0.000	35				
0.05	0.050	0.048	-0.002	239	0.0 %	239				
0.2	0.200	0.200	-0.000	863	0.0 %	863				
2.5	2.500	2.483	-0.017	10225	0.0 %	10225				
5.0	5.000	4.995	-0.005	20529	0.0 %	20529				
7.5	7.500	7.584	0.084	31146	0.0 %	31146				
10.0	10.000	9.943	-0.057	40824	0.0 %	40824				

LB101144 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Mar 2019 14:31:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	35	-	0.0000	245.1	PPB	04 Mar 2019 14:31:49	S	Std
0.05 - 1	50-05	239	-	0.0500	245.1	PPB	04 Mar 2019 14:34:06	S	Std
0.2 - 1	50-2	863	-	0.2000	245.1	PPB	04 Mar 2019 14:36:23	S	Std
2.5 - 1	52-5	10225	-	2.5000	245.1	PPB	04 Mar 2019 14:38:41	S	Std
5.0 - 1	55-0	20529	-	5.0000	245.1	PPB	04 Mar 2019 14:40:57	S	Std
7.5 - 1	57-5	31146	-	7.5000	245.1	PPB	04 Mar 2019 14:43:13	S	Std
10.0 - 1	510-0	40824	-	10.0000	245.1	PPB	04 Mar 2019 14:45:29	S	Std
ICV38 - 1	ICV38	16093	3.9133	-245.1	PPB	PPB	04 Mar 2019 14:48:11	U	SMPL
ICB38 - 1	ICB38	5	-0.0093	-245.1	PPB	PPB	04 Mar 2019 14:50:27	U	SMPL
CCV79 - 1	CCV79	20394	4.9620	-245.1	PPB	PPB	04 Mar 2019 14:52:42	U	SMPL
CCB79 - 1	CCB79	-65	-0.0263	-245.1	PPB	PPB	04 Mar 2019 14:54:57	U	SMPL
CRA - 1	CRA	254	0.0514	-245.1	PPB	PPB	04 Mar 2019 14:57:12	U	SMPL
HighStd - 1	HighStd	41112	10.0134	-245.1	PPB	PPB	04 Mar 2019 14:59:28	U	SMPL
ChkStd - 1	ChkStd	29396	7.1568	-245.1	PPB	PPB	04 Mar 2019 15:01:43	U	SMPL
PB117602BL - 1	PBW	-44	-0.0212	-245.1	PPB	PPB	04 Mar 2019 15:04:00	U	SMPL
PB117602BS - 1	LCSW	16226	3.9457	-245.1	PPB	PPB	04 Mar 2019 15:09:12	U	SMPL
K1717-01 - 1	TANKA_190214	93	0.0122	-245.1	PPB	PPB	04 Mar 2019 15:11:29	U	SMPL
K1717-01DUP - 1	TANKA_190214DUP	123	0.0195	-245.1	PPB	PPB	04 Mar 2019 15:13:46	U	SMPL
K1717-01MS - 1	TANKA_190214MS	16518	4.0169	-245.1	PPB	PPB	04 Mar 2019 15:16:02	U	SMPL
K1717-01MSD - 1	TANKA_190214MSD	16295	3.9626	-245.1	PPB	PPB	04 Mar 2019 15:18:20	U	SMPL
K1717-02 - 1	TANKB_190214	44	0.0002	-245.1	PPB	PPB	04 Mar 2019 15:20:35	U	SMPL
CCV80 - 1	CCV80	20636	5.0210	-245.1	PPB	PPB	04 Mar 2019 15:22:50	U	SMPL
CCB80 - 1	CCB80	19	-0.0058	-245.1	PPB	PPB	04 Mar 2019 15:25:05	U	SMPL
K1717-03 - 1	TANKC_190214	125	0.0200	-245.1	PPB	PPB	04 Mar 2019 15:27:21	U	SMPL
K1717-04 - 1	TANKA_190214	130	0.0212	-245.1	PPB	PPB	04 Mar 2019 15:29:37	U	SMPL
K1717-05 - 1	TANKB_190214	124	0.0198	-245.1	PPB	PPB	04 Mar 2019 15:31:52	U	SMPL
K1717-06 - 1	TANKC_190214	72	0.0071	-245.1	PPB	PPB	04 Mar 2019 15:34:08	U	SMPL
K1717-01LX5 - 1		32	-0.0027	-245.1	PPB	PPB	04 Mar 2019 15:36:24	U	SMPL
K1717-01A - 1		16367	3.9801	-245.1	PPB	PPB	04 Mar 2019 15:38:41	U	SMPL
CCV81 - 1	CCV81	20704	5.0376	-245.1	PPB	PPB	04 Mar 2019 15:40:57	U	SMPL
CCB81 - 1	CCB81	14	-0.0071	-245.1	PPB	PPB	04 Mar 2019 15:43:14	U	SMPL