

LB101146


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58819



A= 0.0000e+000

B= 1.6946e-004

C= 1.4155e-002

Rho= 0.9999316

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	12	0.000	12				
0.2	0.200	0.222	0.022	1224	0.0 %	1224				
2.5	2.500	2.467	-0.033	14477	0.0 %	14477				
5.0	5.000	4.940	-0.060	29066	0.0 %	29066				
7.5	7.500	7.574	0.074	44609	0.0 %	44609				
10.0	10.000	9.982	-0.018	58819	0.0 %	58819				

LB101146 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Mar 2019 15:53:35

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	12	-	0.0000	7471B	PPB	04 Mar 2019 15:53:47	S	Std
0.2 - 1	50-2	1224	-	0.2000	7471B	PPB	04 Mar 2019 15:56:05	S	Std
2.5 - 1	52-5	14477	-	2.5000	7471B	PPB	04 Mar 2019 15:58:22	S	Std
5.0 - 1	55-0	29066	-	5.0000	7471B	PPB	04 Mar 2019 16:00:39	S	Std
7.5 - 1	57-5	44609	-	7.5000	7471B	PPB	04 Mar 2019 16:02:56	S	Std
10.0 - 1	510-0	58819	-	10.0000	7471B	PPB	04 Mar 2019 16:05:13	S	Std
ICV39 - 1	ICV39	22875	3.8905	-	7471B	PPB	04 Mar 2019 16:08:02	U	SMPL
ICB39 - 1	ICB39	67	0.0255	-	7471B	PPB	04 Mar 2019 16:10:18	U	SMPL
CCV82 - 1	CCV82	28126	4.7804	-	7471B	PPB	04 Mar 2019 16:12:34	U	SMPL
CCB82 - 1	CCB82	-76	0.0013	-	7471B	PPB	04 Mar 2019 16:14:49	U	SMPL
CRA - 1	CRA	1198	0.2172	-	7471B	PPB	04 Mar 2019 16:17:06	U	SMPL
HighStd - 1	HighStd	58867	9.9897	-	7471B	PPB	04 Mar 2019 16:19:21	U	SMPL
ChkStd - 1	ChkStd	41607	7.0649	-	7471B	PPB	04 Mar 2019 16:21:36	U	SMPL
PB117584BL - 1	PBS	119	0.0343	-	7471B	PPB	04 Mar 2019 16:23:54	U	SMPL
PB117584BS - 1	LCSS	20709	3.5235	-	7471B	PPB	04 Mar 2019 16:26:11	U	SMPL
K1688-01 - 1	JC-03-030119-A	14594	2.4872	-	7471B	PPB	04 Mar 2019 16:28:28	U	SMPL
K1688-01DUP - 1	JC-03-030119-ADUP	14595	2.4874	-	7471B	PPB	04 Mar 2019 16:30:46	U	SMPL
K1688-01MS - 1	JC-03-030119-AMS	37183	6.3152	-	7471B	PPB	04 Mar 2019 16:33:02	U	SMPL
K1688-01MSD - 1	JC-03-030119-AMSD	36644	6.2238	-	7471B	PPB	04 Mar 2019 16:35:20	U	SMPL
K1688-03 - 1	JC-03-030119-B	8722	1.4922	-	7471B	PPB	04 Mar 2019 16:37:37	U	SMPL
CCV83 - 1	CCV83	27427	4.6619	-	7471B	PPB	04 Mar 2019 16:39:53	U	SMPL
CCB83 - 1	CCB83	-87	-0.0006	-	7471B	PPB	04 Mar 2019 16:42:09	U	SMPL
K1688-05 - 1	JC-03-030119-C	22723	3.8648	-	7471B	PPB	04 Mar 2019 16:44:27	U	SMPL
K1688-07 - 1	JC-03-030119-D	13825	2.3569	-	7471B	PPB	04 Mar 2019 16:46:42	U	SMPL
K1688-09 - 1	JC-03-030119-E	13776	2.3486	-	7471B	PPB	04 Mar 2019 16:48:58	U	SMPL
K1688-11 - 1	JC-03-030119-F	4933	0.8501	-	7471B	PPB	04 Mar 2019 16:51:14	U	SMPL
K1688-13 - 1	JC-03-030119-G	28938	4.9180	-	7471B	PPB	04 Mar 2019 16:53:30	U	SMPL
K1688-15 - 1	JC-03-030119-H	9178	1.5695	-	7471B	PPB	04 Mar 2019 16:55:46	U	SMPL
K1688-17 - 1	JC-03-030119-I	18746	3.1908	-	7471B	PPB	04 Mar 2019 16:58:04	U	SMPL
K1688-19 - 1	JC-03-030119-J	9854	1.6840	-	7471B	PPB	04 Mar 2019 17:00:21	U	SMPL
K1712-05 - 1	Z-4-2-AB	33258	5.6500	-	7471B	PPB	04 Mar 2019 17:02:38	U	SMPL
K1712-07 - 1	Z-2-1	42842	7.2741	-	7471B	PPB	04 Mar 2019 17:04:54	U	SMPL
CCV84 - 1	CCV84	27340	4.6472	-	7471B	PPB	04 Mar 2019 17:07:11	U	SMPL
CCB84 - 1	CCB84	-74	0.0016	-	7471B	PPB	04 Mar 2019 17:09:29	U	SMPL
K1712-11 - 1	Z-2-2	52236	8.8660	-	7471B	PPB	04 Mar 2019 17:11:46	U	SMPL
K1712-15 - 1	Z-2-3	97387	16.5173	-	7471B	PPB	04 Mar 2019 17:14:01	U	SMPL
K1712-19 - 1	Z-2-4	2531	0.4431	-	7471B	PPB	04 Mar 2019 17:16:18	U	SMPL
K1715-03 - 1	TP-4-B-S	1450	0.2599	-	7471B	PPB	04 Mar 2019 17:18:36	U	SMPL
K1715-07 - 1	TP-4-B-D	3167	0.5508	-	7471B	PPB	04 Mar 2019 17:20:52	U	SMPL
K1715-11 - 1	TP-5-A	2410	0.4226	-	7471B	PPB	04 Mar 2019 17:23:08	U	SMPL
PB117585BL - 1	PBS	-31	0.0089	-	7471B	PPB	04 Mar 2019 17:25:25	U	SMPL
PB117585BS - 1	LCSS	20896	3.5552	-	7471B	PPB	04 Mar 2019 17:27:42	U	SMPL
K1707-01 - 1	S10	41501	7.0469	-	7471B	PPB	04 Mar 2019 17:29:59	U	SMPL
K1707-01DUP - 1	S10DUP	41532	7.0521	-	7471B	PPB	04 Mar 2019 17:32:15	U	SMPL
CCV85 - 1	CCV85	27606	4.6923	-	7471B	PPB	04 Mar 2019 17:34:32	U	SMPL
CCB85 - 1	CCB85	-126	-0.0072	-	7471B	PPB	04 Mar 2019 17:36:49	U	SMPL
K1707-01MS - 1	S10MS	68200	11.5713	-	7471B	PPB	04 Mar 2019 17:39:07	U	SMPL
K1707-01MSD - 1	S10MSD	67130	11.3900	-	7471B	PPB	04 Mar 2019 17:41:23	U	SMPL
K1707-03 - 1	S11	32830	5.5775	-	7471B	PPB	04 Mar 2019 17:43:41	U	SMPL
K1707-05 - 1	S12	41498	7.0464	-	7471B	PPB	04 Mar 2019 17:45:58	U	SMPL
K1714-01 - 1	SB-A-0-2	719	0.1360	-	7471B	PPB	04 Mar 2019 17:48:15	U	SMPL
K1714-02 - 1	SB-A-2-5	2413	0.4231	-	7471B	PPB	04 Mar 2019 17:50:33	U	SMPL
K1714-03 - 1	SB-A-5-8	3126	0.5439	-	7471B	PPB	04 Mar 2019 17:52:50	U	SMPL
K1714-04 - 1	SB-A-8-11	1180	0.2141	-	7471B	PPB	04 Mar 2019 17:55:07	U	SMPL
K1714-05 - 1	SB-A-11-14	392	0.0806	-	7471B	PPB	04 Mar 2019 17:57:25	U	SMPL
K1714-06 - 1	SB-A-14-18	720	0.1362	-	7471B	PPB	04 Mar 2019 17:59:41	U	SMPL
CCV86 - 1	CCV86	28353	4.8188	-	7471B	PPB	04 Mar 2019 18:01:56	U	SMPL
CCB86 - 1	CCB86	-79	0.0008	-	7471B	PPB	04 Mar 2019 18:04:12	U	SMPL
K1714-07 - 1	SB-B-0-2	906	0.1677	-	7471B	PPB	04 Mar 2019 18:06:29	U	SMPL
K1714-08 - 1	SB-B-2-5	662	0.1263	-	7471B	PPB	04 Mar 2019 18:08:45	U	SMPL
K1714-09 - 1	SB-B-5-8	893	0.1655	-	7471B	PPB	04 Mar 2019 18:11:01	U	SMPL
K1714-10 - 1	SB-B-8-11	669	0.1275	-	7471B	PPB	04 Mar 2019 18:13:17	U	SMPL
K1714-11 - 1	SB-B-11-14	670	0.1277	-	7471B	PPB	04 Mar 2019 18:15:33	U	SMPL
K1714-12 - 1	SB-B-14-18	411	0.0838	-	7471B	PPB	04 Mar 2019 18:17:49	U	SMPL
K1714-13 - 1	SB-C-0-2	793	0.1485	-	7471B	PPB	04 Mar 2019 18:20:05	U	SMPL
K1714-14 - 1	SB-C-2-5	521	0.1024	-	7471B	PPB	04 Mar 2019 18:22:23	U	SMPL
K1714-15 - 1	SB-C-5-8	549	0.1072	-	7471B	PPB	04 Mar 2019 18:24:40	U	SMPL
K1714-16 - 1	SB-D-0-2	661	0.1262	-	7471B	PPB	04 Mar 2019 18:26:56	U	SMPL
CCV87 - 1	CCV87	27979	4.7555	-	7471B	PPB	04 Mar 2019 18:29:12	U	SMPL
CCB87 - 1	CCB87	-65	0.0031	-	7471B	PPB	04 Mar 2019 18:31:28	U	SMPL
K1714-17 - 1	SB-D-2-5	374	0.0775	-	7471B	PPB	04 Mar 2019 18:33:46	U	SMPL
K1714-18 - 1	SB-D-5-8	590	0.1141	-	7471B	PPB	04 Mar 2019 18:36:03	U	SMPL
K1688-01LX5 - 1		2818	0.4917	-	7471B	PPB	04 Mar 2019 18:38:20	U	SMPL
K1688-01A - 1		37283	6.3321	-	7471B	PPB	04 Mar 2019 18:40:38	U	SMPL
K1707-01LX5 - 1		7870	1.3478	-	7471B	PPB	04 Mar 2019 18:42:54	U	SMPL
K1707-01A - 1		68002	11.5377	-	7471B	PPB	04 Mar 2019 18:45:12	U	SMPL
K1712-15DLX2 - 1		48495	8.2321	-	7471B	PPB	04 Mar 2019 18:47:29	U	SMPL
K1707-01DLX2 - 1		19982	3.4003	-	7471B	PPB	04 Mar 2019 18:49:47	U	SMPL
K1707-01DUPDLX2 - 1		20337	3.4605	-	7471B	PPB	04 Mar 2019 18:52:06	U	SMPL
CCV88 - 1	CCV88	28196	4.7922	-	7471B	PPB	04 Mar 2019 18:54:24	U	SMPL
CCB88 - 1	CCB88	-102	-0.0031	-	7471B	PPB	04 Mar 2019 18:56:40	U	SMPL

LB101146 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Mar 2019 15:53:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
K1707-01MSDLX2 - 1		30394	5.1647		-7471B	PPB	04 Mar 2019 18:58:58	U	SMPL
K1707-01MSDDLX2 - 1		29375	4.9920		-7471B	PPB	04 Mar 2019 19:01:14	U	SMPL
K1707-01ADLX2 - 1		30157	5.1245		-7471B	PPB	04 Mar 2019 19:03:31	U	SMPL
CCV89 - 1	CCV89	28121	4.7795		-7471B	PPB	04 Mar 2019 19:05:47	U	SMPL
CCB89 - 1	CCB89	-89	-0.0009		-7471B	PPB	04 Mar 2019 19:08:05	U	SMPL