



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101167

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	03/05/2019	11:10
2	CAL2	1	Water	NA	NA	21.1	7.00	03/05/2019	11:11
3	CAL3	1	Water	NA	NA	21.1	10.02	03/05/2019	11:12
4	ICV	1	Water	NA	NA	21.2	7.00	03/05/2019	11:15
5	CCV1	1	Water	NA	NA	21.1	2.01	03/05/2019	11:16
6	K1737-01	1	Water	NA	NA	14.7	4.51	03/05/2019	11:18
7	K1737-01DUP	1	Water	NA	NA	14.8	4.52	03/05/2019	11:19
8	CCV2	1	Water	NA	NA	21.2	12.01	03/05/2019	11:20

LP 101167

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1737

WorkList ID : 123025

Date : 03-05-2019 10:55:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2019	Water	K1737-01-17	pH	Cool 4 deg C	STER04	P31	POOL-WATER	03/04/2019	SM 4500-H B

Date/Time 03-05-19 11:00
Received by: Jd
Relinquished by: Cd

Date/Time 03-05-19 11:35
Received by: Cd
Relinquished by: Jd