


LB101169

7470A

INSTRUMENT ID : CUI

 μ Abs. :

41585



Linear

A= 0.0000e+000

B= 2.4037e-004

C= -2.2987e-002

Rho= 0.9999430

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	31	0.000	31				
0.2	0.200	0.213	0.013	982	0.0 %	982				
2.5	2.500	2.522	0.022	10586	0.0 %	10586				
5.0	5.000	4.942	-0.058	20655	0.0 %	20655				
7.5	7.500	7.566	0.066	31571	0.0 %	31571				
10.0	10.000	9.973	-0.027	41585	0.0 %	41585				

LB101169 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Mar 2019 14:05:38

Sample ID	Extended ID	µ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50.0	31	-	0.0000	7470A	PPB	05 Mar 2019 14:06:59	S	Std
0.2 - 1	50.2	982	-	0.2000	7470A	PPB	05 Mar 2019 14:09:07	S	Std
2.5 - 1	52.5	10586	-	2.5000	7470A	PPB	05 Mar 2019 14:11:15	S	Std
5.0 - 1	55.0	20655	-	5.0000	7470A	PPB	05 Mar 2019 14:13:23	S	Std
7.5 - 1	57.5	31571	-	7.5000	7470A	PPB	05 Mar 2019 14:15:30	S	Std
10.0 - 1	510.0	41585	-	10.0000	7470A	PPB	05 Mar 2019 14:17:37	S	Std
ICV41 - 1	ICV41	16044	3.8336	-	7470A	PPB	05 Mar 2019 14:21:12	U	SMPL
ICB41 - 1	ICB41	59	-0.0088	-	7470A	PPB	05 Mar 2019 14:23:18	U	SMPL
CCV93 - 1	CCV93	20448	4.8922	-	7470A	PPB	05 Mar 2019 14:25:24	U	SMPL
CCB93 - 1	CCB93	-36	-0.0316	-	7470A	PPB	05 Mar 2019 14:27:30	U	SMPL
CRA - 1	CRA	975	0.2114	-	7470A	PPB	05 Mar 2019 14:29:36	U	SMPL
HighStd - 1	HighStd	40962	9.8233	-	7470A	PPB	05 Mar 2019 14:31:42	U	SMPL
ChkStd - 1	ChkStd	29132	6.9796	-	7470A	PPB	05 Mar 2019 14:33:48	U	SMPL
PB117626BL - 1	PBW	45	-0.0122	-	7470A	PPB	05 Mar 2019 14:35:56	U	SMPL
PB117626BS - 1	LCSW	18095	4.3266	-	7470A	PPB	05 Mar 2019 14:38:03	U	SMPL
K1706-01 - 1	DRUM-1-6	308	0.0510	-	7470A	PPB	05 Mar 2019 14:40:10	U	SMPL
K1707-02 - 1	S10	169	0.0176	-	7470A	PPB	05 Mar 2019 14:42:18	U	SMPL
K1707-04 - 1	S11	123	0.0066	-	7470A	PPB	05 Mar 2019 14:44:26	U	SMPL
K1707-06 - 1	S12	128	0.0078	-	7470A	PPB	05 Mar 2019 14:46:34	U	SMPL
K1709-01 - 1	COR-21816	150	0.0131	-	7470A	PPB	05 Mar 2019 14:48:40	U	SMPL
CCV94 - 1	CCV94	20650	4.9408	-	7470A	PPB	05 Mar 2019 14:50:46	U	SMPL
CCB94 - 1	CCB94	-52	-0.0355	-	7470A	PPB	05 Mar 2019 14:52:52	U	SMPL
K1710-01 - 1	ETC-21642	150	0.0131	-	7470A	PPB	05 Mar 2019 14:54:59	U	SMPL
K1712-06 - 1	Z-4-2-AB	155	0.0143	-	7470A	PPB	05 Mar 2019 14:57:04	U	SMPL
K1712-10 - 1	Z-2-1	268	0.0414	-	7470A	PPB	05 Mar 2019 14:59:11	U	SMPL
K1712-14 - 1	Z-2-2	183	0.0210	-	7470A	PPB	05 Mar 2019 15:01:18	U	SMPL
K1712-18 - 1	Z-2-3	116	0.0049	-	7470A	PPB	05 Mar 2019 15:03:24	U	SMPL
K1712-22 - 1	Z-2-4	137	0.0099	-	7470A	PPB	05 Mar 2019 15:05:32	U	SMPL
K1715-04 - 1	TP-4-B-S	138	0.0102	-	7470A	PPB	05 Mar 2019 15:07:40	U	SMPL
K1715-08 - 1	TP-4-B-D	148	0.0126	-	7470A	PPB	05 Mar 2019 15:09:48	U	SMPL
K1715-12 - 1	TP-5-A	170	0.0179	-	7470A	PPB	05 Mar 2019 15:11:56	U	SMPL
K1715-12DUP - 1	TP-5-ADUP	172	0.0184	-	7470A	PPB	05 Mar 2019 15:14:03	U	SMPL
CCV95 - 1	CCV95	20173	4.8261	-	7470A	PPB	05 Mar 2019 15:16:09	U	SMPL
CCB95 - 1	CCB95	61	-0.0083	-	7470A	PPB	05 Mar 2019 15:18:16	U	SMPL
K1715-12MS - 1	TP-5-AMS	17155	4.1006	-	7470A	PPB	05 Mar 2019 15:20:22	U	SMPL
K1715-12MSD - 1	TP-5-AMSD	16851	4.0276	-	7470A	PPB	05 Mar 2019 15:22:29	U	SMPL
PB117627BL - 1	PBW	96	0.0001	-	7470A	PPB	05 Mar 2019 15:24:35	U	SMPL
PB117627BS - 1	LCSW	16803	4.0160	-	7470A	PPB	05 Mar 2019 15:26:42	U	SMPL
K1688-02 - 1	JC-03-030119-A	56	-0.0095	-	7470A	PPB	05 Mar 2019 15:28:49	U	SMPL
K1688-02DUP - 1	JC-03-030119-ADUP	122	0.0063	-	7470A	PPB	05 Mar 2019 15:30:58	U	SMPL
K1688-02MS - 1	JC-03-030119-AMS	17149	4.0992	-	7470A	PPB	05 Mar 2019 15:33:05	U	SMPL
K1688-02MSD - 1	JC-03-030119-AMSD	18297	4.3752	-	7470A	PPB	05 Mar 2019 15:35:13	U	SMPL
K1688-04 - 1	JC-03-030119-B	95	-0.0002	-	7470A	PPB	05 Mar 2019 15:37:21	U	SMPL
K1688-06 - 1	JC-03-030119-C	140	0.0107	-	7470A	PPB	05 Mar 2019 15:39:29	U	SMPL
CCV96 - 1	CCV96	21028	5.0316	-	7470A	PPB	05 Mar 2019 15:41:38	U	SMPL
CCB96 - 1	CCB96	47	-0.0117	-	7470A	PPB	05 Mar 2019 15:43:44	U	SMPL
K1688-08 - 1	JC-03-030119-D	133	0.0090	-	7470A	PPB	05 Mar 2019 15:45:51	U	SMPL
K1688-10 - 1	JC-03-030119-E	147	0.0123	-	7470A	PPB	05 Mar 2019 15:48:00	U	SMPL
K1688-12 - 1	JC-03-030119-F	143	0.0114	-	7470A	PPB	05 Mar 2019 15:50:07	U	SMPL
K1688-14 - 1	JC-03-030119-G	110	0.0035	-	7470A	PPB	05 Mar 2019 15:52:13	U	SMPL
K1688-16 - 1	JC-03-030119-H	113	0.0042	-	7470A	PPB	05 Mar 2019 15:54:20	U	SMPL
K1688-18 - 1	JC-03-030119-I	110	0.0035	-	7470A	PPB	05 Mar 2019 15:56:27	U	SMPL
K1688-20 - 1	JC-03-030119-J	93	-0.0006	-	7470A	PPB	05 Mar 2019 15:58:35	U	SMPL
PB117628BL - 1	PBW	137	0.0099	-	7470A	PPB	05 Mar 2019 16:00:42	U	SMPL
PB117628BS - 1	LCSW	17399	4.1593	-	7470A	PPB	05 Mar 2019 16:08:45	U	SMPL
K1700-02 - 1	NB-21841-21842	87	-0.0021	-	7470A	PPB	05 Mar 2019 16:10:51	U	SMPL
CCV97 - 1	CCV97	20924	5.0066	-	7470A	PPB	05 Mar 2019 16:12:58	U	SMPL
CCB97 - 1	CCB97	-69	-0.0396	-	7470A	PPB	05 Mar 2019 16:15:05	U	SMPL
K1715-14 - 1	TP-4-B	112	0.0039	-	7470A	PPB	05 Mar 2019 16:17:11	U	SMPL
K1715-14DUP - 1	TP-4-BDUP	141	0.0109	-	7470A	PPB	05 Mar 2019 16:19:18	U	SMPL
K1715-14MS - 1	TP-4-BMS	16729	3.9982	-	7470A	PPB	05 Mar 2019 16:21:24	U	SMPL
K1715-14MSD - 1	TP-4-BMSD	16439	3.9285	-	7470A	PPB	05 Mar 2019 16:23:30	U	SMPL
PB117629BL - 1	PBW	16	-0.0191	-	7470A	PPB	05 Mar 2019 16:25:38	U	SMPL
PB117629BS - 1	LCSW	19004	4.5451	-	7470A	PPB	05 Mar 2019 16:30:56	U	SMPL
K1727-01 - 1	MW-11	659	0.1354	-	7470A	PPB	05 Mar 2019 16:33:03	U	SMPL
K1727-01DUP - 1	MW-11DUP	716	0.1491	-	7470A	PPB	05 Mar 2019 16:35:10	U	SMPL
K1727-01MS - 1	MW-11MS	15028	3.5894	-	7470A	PPB	05 Mar 2019 16:37:18	U	SMPL
K1727-01MSD - 1	MW-11MSD	14745	3.5213	-	7470A	PPB	05 Mar 2019 16:39:25	U	SMPL
CCV98 - 1	CCV98	20890	4.9984	-	7470A	PPB	05 Mar 2019 16:41:33	U	SMPL
CCB98 - 1	CCB98	13	-0.0199	-	7470A	PPB	05 Mar 2019 16:43:42	U	SMPL
K1727-02 - 1	TW-1	607	0.1229	-	7470A	PPB	05 Mar 2019 16:45:50	U	SMPL
K1727-03 - 1	MW-9	175	0.0191	-	7470A	PPB	05 Mar 2019 16:47:58	U	SMPL
K1727-04 - 1	MW-8	327	0.0556	-	7470A	PPB	05 Mar 2019 16:50:06	U	SMPL
K1727-05 - 1	MW-7	697	0.1446	-	7470A	PPB	05 Mar 2019 16:52:14	U	SMPL
K1727-06 - 1	MW-6	643	0.1316	-	7470A	PPB	05 Mar 2019 16:54:23	U	SMPL
K1727-07 - 1	FIELD-BLANK	57	-0.0093	-	7470A	PPB	05 Mar 2019 16:56:30	U	SMPL
K1727-08 - 1	DUP-1	721	0.1503	-	7470A	PPB	05 Mar 2019 16:58:38	U	SMPL
PB117626TB - 1	PB117548TB	127	0.0075	-	7470A	PPB	05 Mar 2019 17:00:46	U	SMPL
PB117627TB - 1	PB117556TB	152	0.0135	-	7470A	PPB	05 Mar 2019 17:02:54	U	SMPL
PB117628TB - 1	PB117550TB	154	0.0140	-	7470A	PPB	05 Mar 2019 17:05:01	U	SMPL
CCV99 - 1	CCV99	20172	4.8259	-	7470A	PPB	05 Mar 2019 17:07:09	U	SMPL

LB101169 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Mar 2019 14:05:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB99 - 1	CCB99	-50	-0.0350	-7470A	PPB		05 Mar 2019 17:09:16	U	SMPL
K1715-12LX5 - 1		26	-0.0167	-7470A	PPB		05 Mar 2019 17:11:24	U	SMPL
K1715-12A - 1		16848	4.0268	-7470A	PPB		05 Mar 2019 17:13:31	U	SMPL
K1688-02LX5 - 1		-9	-0.0252	-7470A	PPB		05 Mar 2019 17:15:38	U	SMPL
K1688-02A - 1		17405	4.1607	-7470A	PPB		05 Mar 2019 17:17:46	U	SMPL
K1715-14LX5 - 1		-33	-0.0309	-7470A	PPB		05 Mar 2019 17:19:53	U	SMPL
K1715-14A - 1		16646	3.9783	-7470A	PPB		05 Mar 2019 17:22:00	U	SMPL
K1727-01LX5 - 1		101	0.0013	-7470A	PPB		05 Mar 2019 17:24:08	U	SMPL
K1727-01A - 1		14917	3.5627	-7470A	PPB		05 Mar 2019 17:26:16	U	SMPL
CCV01 - 1	CCV01	20814	4.9802	-7470A	PPB		05 Mar 2019 17:28:24	U	SMPL
CCB01 - 1	CCB01	-24	-0.0288	-7470A	PPB		05 Mar 2019 17:30:32	U	SMPL