



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/6/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 18:35
In Date: 03/05/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 03/06/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101178

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1732-01	SB-07 (22-23) -20190301	1	1.13	9.77	8.98	90.9
K1732-02	K1732-01MS	2	1.13	9.77	8.98	90.9
K1732-03	K1732-01MSD	3	1.13	9.77	8.98	90.9
K1732-04	SB-107 (22-23) -20190301	4	1.16	9.81	9.09	91.7
K1733-01	ADELINE-ST-CONCRETE	14	1.00	2.00	2.00	100.0
K1733-02	LIBERTY-ST-CONCRETE	15	1.00	2.00	2.00	100.0
K1738-01	ODD	16	1.00	2.00	2.00	100.0
K1738-02	OILY-DEBRIS	17	1.00	2.00	2.00	100.0
K1740-01	COR-INSULATING-OIL	18	1.00	2.00	2.00	100.0
K1741-01	390220727	5	1.18	9.98	1.37	2.2
K1742-01	390226041-DIGESTER	6	1.15	9.7	1.31	1.9
K1743-01	390214016.2-SLUDGE-OUT	7	1.19	9.87	1.41	2.5
K1744-01	390215016	8	1.11	9.56	1.46	4.1
K1745-01	390220710-SQARSLUDGE	9	1.15	9.81	2.96	20.9
K1746-01	390226074-SQAR-CAKE	10	1.14	9.63	2.82	19.8
K1747-01	390207045-BFPSC-020719	11	1.15	9.97	2.5	15.3
K1748-01	390213018	12	1.14	9.94	1.68	6.1
K1752-01	DR-FC	13	1.19	9.98	8.1	78.6
K1753-01	Z-2-5	19	1.13	9.72	8.43	85.0
K1753-02	Z-2-5-A	20	1.12	9.84	8.54	85.1
K1753-03	Z-2-5-B	21	1.12	9.84	8.54	85.1
K1753-04	Z-2-5	22	1.13	9.72	8.43	85.0
K1753-05	Z-2-6	23	1.12	9.68	8.27	83.5
K1753-06	Z-2-6-A	24	1.1	99.76	8.53	7.5
K1753-07	Z-2-6-B	25	1.1	9.76	8.53	85.8
K1753-08	Z-2-6	26	1.12	9.68	8.27	83.5
K1753-09	Z-2-7	27	1.1	9.77	8.33	83.4
K1753-10	Z-2-7-A	28	1.11	9.98	9.00	89.0
K1753-11	Z-2-7-B	29	1.11	9.98	9.00	89.0
K1753-12	Z-2-7	30	1.1	9.77	8.33	83.4

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$