



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101206

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/06/2019	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/06/2019	16:01
3	CAL3	1	Water	NA	NA	20.2	10.02	03/06/2019	16:03
4	ICV	1	Water	NA	NA	20.2	7.01	03/06/2019	16:06
5	CCV1	1	Water	NA	NA	20.1	2.01	03/06/2019	16:06
6	K1761-02	1	Water	NA	NA	12.2	6.90	03/06/2019	16:07
7	K1761-02DUP	1	Water	NA	NA	12.3	6.91	03/06/2019	16:08
8	CCV2	1	Water	NA	NA	20.2	12.01	03/06/2019	16:10

13101206

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K1761

WorkList ID : 123077

Date : 03-06-2019 15:42:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/06/2019	Water	K1761-02 - 17	Corrosivity	Cool 4 deg C	PSEG01	K31	N4871	03/06/2019	9040C

Date/Time 03-06-19 15:50
 Received by: JP
 Relinquished by: Ca

Date/Time 03-06-19 16:15
 Received by: JP
 Relinquished by: Ca