

LB101221

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/7/2019 11:13

Test: Phenolics-

Sample Id	Result ^{mg/l}	Dil. 1 +	Response	Errors
ICV1	1.005	0.0	0.124	
ICB1	0.002	0.0	0.016	
CCV1	0.997	0.0	0.123	
CCB1	0.005	0.0	0.017	
PB117642BL	0.005	0.0	0.016	
PB117642BS	1.027	0.0	0.126	
K1560-03	0.015	0.0	0.018	
PB117644BL	0.000	0.0	0.016	
PB117644BS	1.009	0.0	0.125	
K1560-09	0.015	0.0	0.018	
CCV2	1.002	0.0	0.124	
CCB2	-0.002	0.0	0.016	

N	12
Mean	0.423
SD	0.5161
CV%	121.89

Aquakem v. 7.2AQ1

AS

Results from time period:

Thu Mar 07 10:23:49 2019

Thu Mar 07 11:10:45 2019

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0505	mg/l	3/7/2019 10:23:49	
0.1PPM	A	Phenolics-	P	0.0999	mg/l	3/7/2019 10:23:50	
0.25PPM	A	Phenolics-	P	0.2456	mg/l	3/7/2019 10:23:51	
0.50PPM	A	Phenolics-	P	0.5023	mg/l	3/7/2019 10:23:52	
1.0PPM	A	Phenolics-	P	1.0034	mg/l	3/7/2019 10:23:53	
2.0PPM	A	Phenolics-	P	1.9983	mg/l	3/7/2019 10:23:54	
ICV1	S	Phenolics-	P	1.0051	mg/l	3/7/2019 11:06:46	
ICB1	S	Phenolics-	P	0.0019	mg/l	3/7/2019 11:06:47	
CCV1	S	Phenolics-	P	0.9973	mg/l	3/7/2019 11:06:48	
CCB1	S	Phenolics-	P	0.0055	mg/l	3/7/2019 11:06:49	
PB117642BL	S	Phenolics-	P	0.0046	mg/l	3/7/2019 11:06:50	
PB117642BS	S	Phenolics-	P	1.0267	mg/l	3/7/2019 11:06:51	
K1560-03	S	Phenolics-	P	0.0154	mg/l	3/7/2019 11:06:52	
PB117644BL	S	Phenolics-	P	0	mg/l	3/7/2019 11:06:53	
PB117644BS	S	Phenolics-	P	1.0088	mg/l	3/7/2019 11:06:54	
K1560-09	S	Phenolics-	P	0.0154	mg/l	3/7/2019 11:06:55	
CCV2	S	Phenolics-	P	1.0019	mg/l	3/7/2019 11:10:44	
CCB2	S	Phenolics-	P	-0.002	mg/l	3/7/2019 11:10:45	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/7/2019 10:24


Test Phenolics-

Accepted 3/7/2019 10:24

Factor 9.29
Bias 0.016

Coeff. of det. 0.999986

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.021	0.0505	0.0500	
2	Phenol-2	0.027	0.0999	0.1000	
3	Phenol-2	0.042	0.2456	0.2500	
4	Phenol-2	0.070	0.5023	0.5000	
5	Phenol-2	0.124	1.0034	1.0000	
6	Phenol-2	0.231	1.9983	2.0000	