

LB101222

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/7/2019 11:50

Test: Phenolics-

Sample Id	Result <i>mg/d</i>	Dil. 1 +	Response	Errors
ICV1	1.006	0.0	0.124	
ICB1	0.002	0.0	0.016	
CCV1	0.999	0.0	0.123	
CCB1	-0.010	0.0	0.015	
PB117643BL	0.000	0.0	0.016	
PB117643BS	1.013	0.0	0.125	
K1560-12	0.011	0.0	0.017	
CCV2	1.015	0.0	0.125	
CCB2	-0.009	0.0	0.015	
N	9			
Mean	0.448			
SD	0.5321			
CV%	118.90			

Aquakem v. 7.2AQ1

Results from time period:

Thu Mar 07 11:43:38 2019

Thu Mar 07 11:43:48 2019

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0505	mg/l	3/7/2019 10:23:49	
0.1PPM	A	Phenolics-	P	0.0999	mg/l	3/7/2019 10:23:50	
0.25PPM	A	Phenolics-	P	0.2456	mg/l	3/7/2019 10:23:51	
0.50PPM	A	Phenolics-	P	0.5023	mg/l	3/7/2019 10:23:52	
1.0PPM	A	Phenolics-	P	1.0034	mg/l	3/7/2019 10:23:53	
2.0PPM	A	Phenolics-	P	1.9983	mg/l	3/7/2019 10:23:54	
ICV1	S	Phenolics-	P	1.0064	mg/l	3/7/2019 11:43:38	
ICB1	S	Phenolics-	P	0.0022	mg/l	3/7/2019 11:43:39	
CCV1	S	Phenolics-	P	0.9985	mg/l	3/7/2019 11:43:40	
CCB1	S	Phenolics-	P	-0.0103	mg/l	3/7/2019 11:43:41	
PB117643BL	S	Phenolics-	P	0.0004	mg/l	3/7/2019 11:43:42	
PB117643BS	S	Phenolics-	P	1.0128	mg/l	3/7/2019 11:43:43	
K1560-12	S	Phenolics-	P	0.0112	mg/l	3/7/2019 11:43:44	
CCV2	S	Phenolics-	P	1.0155	mg/l	3/7/2019 11:43:47	
CCB2	S	Phenolics-	P	-0.0091	mg/l	3/7/2019 11:43:48	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/7/2019 10:24


Test Phenolics-

Accepted 3/7/2019 10:24

Factor 9.29
Bias 0.016

Coeff. of det. 0.999986

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.021	0.0505	0.0500	
2	Phenol-2	0.027	0.0999	0.1000	
3	Phenol-2	0.042	0.2456	0.2500	
4	Phenol-2	0.070	0.5023	0.5000	
5	Phenol-2	0.124	1.0034	1.0000	
6	Phenol-2	0.231	1.9983	2.0000	