



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101224

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 8                 | 27.0       | 80.6      | 80.6            | 03/07/2019 | 09:40     |
| 2   | K1761-02    |               | 1  | 9                 | >100.0     | >212.0    | >212.0          | 03/07/2019 | 10:10     |
| 3   | K1761-02DUP |               | 1  | 7                 | >100.0     | >212.0    | >212.0          | 03/07/2019 | 10:35     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB 101224

WorkList Name : FP3719

WorkList ID : 123114

Date : 03-07-2019 08:58:07

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 03/11/2019 | Water  | K1761-02 | Flash Point | Cool 4 deg C | PSEG01   | K31              | N4871           | 03/06/2019   | 1010A  |

Date/Time 03/07/19 09:10

Received by: KS

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

03/07/19 10:40

CP

KS