

LB101231


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

42278



A= 0.0000e+000

B= 2.3630e-004

C= -1.6332e-002

Rho= 0.9999408

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	25	0.000	25				
0.2	0.200	0.178	-0.022	823	0.0 %	823				
2.5	2.500	2.492	-0.008	10616	0.0 %	10616				
5.0	5.000	5.089	0.089	21603	0.0 %	21603				
7.5	7.500	7.477	-0.023	31712	0.0 %	31712				
10.0	10.000	9.974	-0.026	42278	0.0 %	42278				

LB101231 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Mar 2019 13:37:58

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	25	-	0.0000	7470A	PPB	07 Mar 2019 13:41:20	S	Std
0.2 - 1	50.2	823	-	0.2000	7470A	PPB	07 Mar 2019 13:43:28	S	Std
2.5 - 1	52.5	10616	-	2.5000	7470A	PPB	07 Mar 2019 13:45:36	S	Std
5.0 - 1	55.0	21603	-	5.0000	7470A	PPB	07 Mar 2019 13:47:44	S	Std
7.5 - 1	57.5	31712	-	7.5000	7470A	PPB	07 Mar 2019 13:49:52	S	Std
10.0 - 1	510.0	42278	-	10.0000	7470A	PPB	07 Mar 2019 13:51:58	S	Std
ICV44 - 1	ICV44	16184	3.8080	-	7470A	PPB	07 Mar 2019 13:54:38	U	SMPL
ICB44 - 1	ICB44	-42	-0.0263	-	7470A	PPB	07 Mar 2019 13:56:43	U	SMPL
CCV08 - 1	CCV08	21906	5.1601	-	7470A	PPB	07 Mar 2019 13:58:49	U	SMPL
CCB08 - 1	CCB08	-49	-0.0279	-	7470A	PPB	07 Mar 2019 14:00:55	U	SMPL
CRA - 1	CRA	800	0.1727	-	7470A	PPB	07 Mar 2019 14:03:01	U	SMPL
HighStd - 1	HighStd	42558	10.0403	-	7470A	PPB	07 Mar 2019 14:05:07	U	SMPL
ChkStd - 1	ChkStd	29611	6.9809	-	7470A	PPB	07 Mar 2019 14:07:13	U	SMPL
PB117697BL - 1	PBW	-37	-0.0251	-	7470A	PPB	07 Mar 2019 14:09:21	U	SMPL
PB117697BS - 1	LCSW	16905	3.9784	-	7470A	PPB	07 Mar 2019 14:11:28	U	SMPL
K1738-03 - 1	OILY-DEBRIS	-3	-0.0170	-	7470A	PPB	07 Mar 2019 14:13:35	U	SMPL
K1753-04 - 1	Z-2-5	37	-0.0076	-	7470A	PPB	07 Mar 2019 14:15:43	U	SMPL
K1753-04DUP - 1	Z-2-5DUP	49	-0.0048	-	7470A	PPB	07 Mar 2019 14:17:50	U	SMPL
K1753-04MS - 1	Z-2-5MS	16731	3.9373	-	7470A	PPB	07 Mar 2019 14:19:58	U	SMPL
K1753-04MSD - 1	Z-2-5MSD	16564	3.8978	-	7470A	PPB	07 Mar 2019 14:22:05	U	SMPL
CCV09 - 1	CCV09	21474	5.0581	-	7470A	PPB	07 Mar 2019 14:24:11	U	SMPL
CCB09 - 1	CCB09	-40	-0.0258	-	7470A	PPB	07 Mar 2019 14:26:17	U	SMPL
K1753-08 - 1	Z-2-6	44	-0.0059	-	7470A	PPB	07 Mar 2019 14:28:23	U	SMPL
K1753-12 - 1	Z-2-7	41	-0.0066	-	7470A	PPB	07 Mar 2019 14:30:29	U	SMPL
K1754-01 - 1	PLA-DEBRIS	273	0.0482	-	7470A	PPB	07 Mar 2019 14:32:36	U	SMPL
PB117698BL - 1	PBW	41	-0.0066	-	7470A	PPB	07 Mar 2019 14:34:42	U	SMPL
PB117698BS - 1	LCSW	16910	3.9796	-	7470A	PPB	07 Mar 2019 14:36:49	U	SMPL
K1754-02 - 1	PLA-21835	12	-0.0135	-	7470A	PPB	07 Mar 2019 14:38:56	U	SMPL
K1754-02DUP - 1	PLA-21835DUP	17	-0.0123	-	7470A	PPB	07 Mar 2019 14:41:03	U	SMPL
K1754-02MS - 1	PLA-21835MS	17331	4.0791	-	7470A	PPB	07 Mar 2019 14:43:11	U	SMPL
K1754-02MSD - 1	PLA-21835MSD	17010	4.0032	-	7470A	PPB	07 Mar 2019 14:45:20	U	SMPL
K1755-25 - 1	COMP-1	-11	-0.0189	-	7470A	PPB	07 Mar 2019 14:47:26	U	SMPL
CCV10 - 1	CCV10	21733	5.1193	-	7470A	PPB	07 Mar 2019 14:49:34	U	SMPL
CCB10 - 1	CCB10	-48	-0.0277	-	7470A	PPB	07 Mar 2019 14:51:40	U	SMPL
K1755-26 - 1	COMP-2	18	-0.0121	-	7470A	PPB	07 Mar 2019 14:53:47	U	SMPL
K1755-27 - 1	COMP-3	31	-0.0090	-	7470A	PPB	07 Mar 2019 14:55:53	U	SMPL
K1755-28 - 1	COMP-4	59	-0.0024	-	7470A	PPB	07 Mar 2019 14:57:59	U	SMPL
K1755-29 - 1	COMP-5	26	-0.0102	-	7470A	PPB	07 Mar 2019 15:00:06	U	SMPL
PB117699BL - 1	PBW	14	-0.0130	-	7470A	PPB	07 Mar 2019 15:02:13	U	SMPL
PB117699BS - 1	LCSW	16470	3.8756	-	7470A	PPB	07 Mar 2019 15:04:20	U	SMPL
K1761-01 - 1	N4871	50	-0.0045	-	7470A	PPB	07 Mar 2019 15:06:27	U	SMPL
K1761-01DUP - 1	N4871DUP	82	0.0030	-	7470A	PPB	07 Mar 2019 15:08:35	U	SMPL
K1761-01MS - 1	N4871MS	18971	4.4666	-	7470A	PPB	07 Mar 2019 15:10:44	U	SMPL
K1761-01MSD - 1	N4871MSD	19019	4.4779	-	7470A	PPB	07 Mar 2019 15:12:50	U	SMPL
CCV11 - 1	CCV11	20791	4.8967	-	7470A	PPB	07 Mar 2019 15:14:57	U	SMPL
CCB11 - 1	CCB11	-46	-0.0272	-	7470A	PPB	07 Mar 2019 15:17:04	U	SMPL
PB117697TB - 1	PB117650TB	15	-0.0128	-	7470A	PPB	07 Mar 2019 15:19:10	U	SMPL
PB117698TB - 1	PB117651TB	28	-0.0097	-	7470A	PPB	07 Mar 2019 15:21:17	U	SMPL
K1753-04LX5 - 1		7	-0.0147	-	7470A	PPB	07 Mar 2019 15:23:23	U	SMPL
K1753-04A - 1		16611	3.9089	-	7470A	PPB	07 Mar 2019 15:25:30	U	SMPL
K1754-02LX5 - 1		-37	-0.0251	-	7470A	PPB	07 Mar 2019 15:27:37	U	SMPL
K1754-02A - 1		16827	3.9600	-	7470A	PPB	07 Mar 2019 15:29:44	U	SMPL
K1761-01LX5 - 1		-30	-0.0234	-	7470A	PPB	07 Mar 2019 15:31:52	U	SMPL
K1761-01A - 1		18916	4.4536	-	7470A	PPB	07 Mar 2019 15:34:00	U	SMPL
CCV12 - 1	CCV12	21603	5.0885	-	7470A	PPB	07 Mar 2019 15:36:08	U	SMPL
CCB12 - 1	CCB12	-44	-0.0267	-	7470A	PPB	07 Mar 2019 15:38:15	U	SMPL