



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101240

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/07/2019	15:45
2	CAL2	1	Water	NA	NA	20.2	7.01	03/07/2019	15:46
3	CAL3	1	Water	NA	NA	20.2	10.02	03/07/2019	15:47
4	ICV	1	Water	NA	NA	20.2	7.00	03/07/2019	15:48
5	CCV1	1	Water	NA	NA	20.2	2.01	03/07/2019	15:49
6	K1783-01	1	Water	NA	NA	23.2	7.06	03/07/2019	15:50
7	K1783-02	1	Water	NA	NA	22.4	7.01	03/07/2019	15:51
8	K1783-03	1	Water	NA	NA	22.9	7.01	03/07/2019	15:53
9	K1783-04	1	Water	NA	NA	22.6	6.91	03/07/2019	16:00
10	K1783-04DUP	1	Water	NA	NA	22.5	6.92	03/07/2019	16:01
11	K1783-07	1	Water	NA	NA	22.2	6.38	03/07/2019	16:02
12	K1783-08	1	Water	NA	NA	22.2	5.72	03/07/2019	16:05
13	CCV2	1	Water	NA	NA	22.1	12.01	03/07/2019	16:06

LB101240

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph K1783

WorkList ID : 123139

Date : 03-07-2019 15:33:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2019	Water	K1783-01- 1	pH	Cool 4 deg C	WEST10	P41	SW-4	03/05/2019	9040C
03/05/2019	Water	K1783-02- 1	pH	Cool 4 deg C	WEST10	P41	SW-3	03/05/2019	9040C
03/05/2019	Water	K1783-03- 1	pH	Cool 4 deg C	WEST10	P41	2019-03-05-DUP-SW	03/05/2019	9040C
03/05/2019	Water	K1783-04- 1	pH	Cool 4 deg C	WEST10	P41	SW-2	03/05/2019	9040C
03/05/2019	Water	K1783-07- 1	pH	Cool 4 deg C	WEST10	P41	SW-1	03/05/2019	9040C
03/05/2019	Water	K1783-08- 1	pH	Cool 4 deg C	WEST10	P41	FB-2019-03-05	03/05/2019	9040C

Date/Time 03.07.19 16.40
Received by: JP
Relinquished by: CO

Date/Time 03.07.19 16.30
Received by: JP
Relinquished by: CO