



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/8/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:40
In Date: 03/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:33
Out Date: 03/08/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101245

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1693-01	OR-01-030719-A	1	1.14	9.76	8.91	90.1
K1693-03	OR-01-030719-B	2	1.12	9.6	8.97	92.6
K1749-01	390219068-SLUDGE-THICKN	3	1.18	9.98	1.28	1.1
K1787-01	SP-1	4	1.14	9.96	9.41	93.8
K1787-02	SP-1-A	5	1.14	9.96	9.41	93.8
K1787-03	SP-1-B	6	1.14	9.96	9.41	93.8
K1787-04	SP-1	7	1.14	9.96	9.41	93.8
K1797-01	3-7-A	8	1.00	2.00	2.00	100.0
K1797-02	3-7-B	9	1.00	2.00	2.00	100.0
K1807-01	TP-1	10	1.17	9.98	8.97	88.5
K1807-03	TP-2	11	1.16	9.8	8.42	84.0
K1807-05	TP-3	12	1.15	9.58	8.24	84.1
K1807-07	TP-4	13	1.12	9.62	8.69	89.1
K1807-09	TP-5	14	1.14	9.77	8.63	86.8
K1807-11	TP-6	15	1.19	9.68	8.58	87.0
K1807-13	TP-7	16	1.15	9.95	8.94	88.5
K1807-15	TP-8	17	1.18	9.94	8.8	87.0
K1807-17	TP-9	18	1.19	9.86	8.63	85.8
K1807-19	TP-10	19	1.16	9.77	9.00	91.1
K1807-21	TP-18	20	1.15	9.83	9.02	90.7
K1807-23	TP-17	21	1.17	9.81	8.52	85.1
K1807-25	TP-16	22	1.13	9.77	8.67	87.3
K1807-27	TP-15	23	1.1	9.81	8.79	88.3
K1807-29	TP-14	24	1.17	9.59	8.53	87.4
K1807-31	TP-13	25	1.15	9.51	8.49	87.8
K1807-33	TP-12	26	1.16	9.55	8.91	92.4
K1807-35	TP-11	27	1.15	9.94	8.78	86.8
K1807-37	TP-19	28	1.18	9.89	8.8	87.5
K1807-39	TP-20	29	1.16	9.85	8.61	85.7
K1807-41	TP-21	30	1.18	9.98	9.28	92.0
K1807-43	TP-22	31	1.11	9.75	8.72	88.1
K1807-45	TP-23	32	1.12	9.89	8.36	82.6
K1808-01	FK-BPM-01-AA0-BB0-AA0A-1	33	1.14	9.84	9.19	92.5
K1808-02	FK-BPM-01-AA0-BB0-AA0A-1	34	1.13	9.62	9.00	92.7
K1809-01	FK-BPM-01-A0-B0-A0A-A0B	35	1.16	9.6	8.99	92.8
K1809-02	FK-BPM-01-A0-B0-A0A-A0B	36	1.19	9.64	9.03	92.8

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$