



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101246

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/07/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/07/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.02	03/07/2019	18:02
4	ICV	1	Water	NA	NA	20.3	7.00	03/07/2019	18:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/07/2019	18:06
6	K1693-02	1	Solid	20.02	20	21.7	7.22	03/07/2019	18:07
7	K1693-02DUP	1	Solid	20.04	20	21.8	7.23	03/07/2019	18:09
8	K1693-04	1	Solid	20.05	20	19.7	8.16	03/07/2019	18:10
9	K1787-04	1	Solid	20.02	20	19.6	7.85	03/07/2019	18:11
10	CCV2	1	Water	NA	NA	20.3	12.01	03/07/2019	18:13

WORKLIST(Hardcopy Internal Chain)

LB 101246

WorkList Name : CORROSIVITY K1305

WorkList ID : 123148

Date : 03-07-2019 16:38:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/07/2019	Solid	K1693-02	Corrosivity	Cool 4 deg C	PSEG05	K31	OR-01-030719-A	03/07/2019	9045D
03/07/2019	Solid	K1693-04	Corrosivity	Cool 4 deg C	PSEG05	K31	OR-01-030719-B	03/07/2019	9045D
03/07/2019	Solid	K1787-04	Corrosivity	Cool 4 deg C	PSEG01	K31	SP-1	03/07/2019	9045D

Date/Time 03-07-19 17:00
 Received by: Jo
 Relinquished by: CA

Date/Time 03-07-19 18:20
 Received by: SP
 Relinquished by: CA