



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101258

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	03/08/2019	11:30
2	CAL2	1	Water	NA	NA	21.1	7.01	03/08/2019	11:31
3	CAL3	1	Water	NA	NA	21.1	10.02	03/08/2019	11:32
4	ICV	1	Water	NA	NA	21.2	6.99	03/08/2019	11:34
5	CCV1	1	Water	NA	NA	21.1	2.01	03/08/2019	11:35
6	K1820-01	1	Water	NA	NA	17.0	4.06	03/08/2019	11:36
7	K1820-01DUP	1	Water	NA	NA	17.1	4.07	03/08/2019	11:37
8	K1820-02	1	Water	NA	NA	17.8	4.17	03/08/2019	11:39
9	K1820-03	1	Water	NA	NA	17.9	4.29	03/08/2019	11:40
10	CCV2	1	Water	NA	NA	21.1	12.01	03/08/2019	11:42

LB101258

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1820

WorkList ID : 123181

Date : 03-08-2019 11:23:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/07/2019	Water	K1820-01 - 1A	pH	Cool 4 deg C	STER13	P11	CELL-A-POOL-WATER	03/07/2019	SM 4500-H B
03/07/2019	Water	K1820-02 - 1A	pH	Cool 4 deg C	STER13	P11	CELL-B-POOL-WATER	03/07/2019	SM 4500-H B
03/07/2019	Water	K1820-03 - 1A	pH	Cool 4 deg C	STER13	P11	CELL-C-POOL-WATER	03/07/2019	SM 4500-H B

Date/Time 03-08-19 11:25
 Received by: Jp
 Relinquished by: Cg

Date/Time 03-08-19 12:00
 Received by: Jp
 Relinquished by: Cg