


LB101266

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.047	0.047	-6	0.000	-6				
0.2	0.200	0.225	0.025	763	0.0 %	763				
2.5	2.500	2.442	-0.058	10350	0.0 %	10350				
5.0	5.000	4.921	-0.079	21069	0.0 %	21069				
7.5	7.500	7.553	0.053	32449	0.0 %	32449				
10.0	10.000	10.012	0.012	43079	0.0 %	43079				

LB101266 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Mar 2019 14:25:15

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-6	-	0.00007470A	PPB		08 Mar 2019 14:55:55	S	Std
0.2 - 1	50.2	763	-	0.20007470A	PPB		08 Mar 2019 14:58:02	S	Std
2.5 - 1	52.5	10350	-	2.50007470A	PPB		08 Mar 2019 15:00:11	S	Std
5.0 - 1	55.0	21069	-	5.00007470A	PPB		08 Mar 2019 15:02:18	S	Std
7.5 - 1	57.5	32449	-	7.50007470A	PPB		08 Mar 2019 15:04:26	S	Std
10.0 - 1	510.0	43079	-	10.00007470A	PPB		08 Mar 2019 15:09:45	S	Std
ICV48 - 1	ICV48	16620	3.8922	-7470A	PPB		08 Mar 2019 15:12:46	U	SMPL
ICB48 - 1	ICB48	-61	0.0341	-7470A	PPB		08 Mar 2019 15:14:52	U	SMPL
CCV21 - 1	CCV21	21295	4.9735	-7470A	PPB		08 Mar 2019 15:16:58	U	SMPL
CCB21 - 1	CCB21	-41	0.0387	-7470A	PPB		08 Mar 2019 15:19:04	U	SMPL
CRA - 1	CRA	715	0.2136	-7470A	PPB		08 Mar 2019 15:21:09	U	SMPL
HighStd - 1	HighStd	42752	9.9362	-7470A	PPB		08 Mar 2019 15:23:15	U	SMPL
ChkStd - 1	ChkStd	31045	7.2286	-7470A	PPB		08 Mar 2019 15:25:22	U	SMPL
PB117743BL - 1	PBW	-77	0.0304	-7470A	PPB		08 Mar 2019 15:27:30	U	SMPL
PB117743BS - 1	LCSW	17571	4.1122	-7470A	PPB		08 Mar 2019 15:29:37	U	SMPL
K1757-04 - 1	TP-6CD-S	-22	0.0431	-7470A	PPB		08 Mar 2019 15:31:44	U	SMPL
K1757-04DUP - 1	TP-6CD-SDUP	21	0.0531	-7470A	PPB		08 Mar 2019 15:33:52	U	SMPL
K1757-04MS - 1	TP-6CD-SMS	16071	3.7653	-7470A	PPB		08 Mar 2019 15:36:00	U	SMPL
K1757-04MSD - 1	TP-6CD-SMSD	15975	3.7431	-7470A	PPB		08 Mar 2019 15:38:07	U	SMPL
K1757-08 - 1	TP-6CD-D	-28	0.0418	-7470A	PPB		08 Mar 2019 15:40:13	U	SMPL
CCV22 - 1	CCV22	21066	4.9205	-7470A	PPB		08 Mar 2019 15:42:19	U	SMPL
CCB22 - 1	CCB22	-44	0.0381	-7470A	PPB		08 Mar 2019 15:44:25	U	SMPL
K1757-12 - 1	TP-6AB-S	3	0.0489	-7470A	PPB		08 Mar 2019 15:46:31	U	SMPL
K1757-16 - 1	TP-6AB-D	27	0.0545	-7470A	PPB		08 Mar 2019 15:48:38	U	SMPL
K1760-04 - 1	Z2-8	11	0.0508	-7470A	PPB		08 Mar 2019 15:50:44	U	SMPL
K1760-08 - 1	Z2-9	6	0.0496	-7470A	PPB		08 Mar 2019 15:52:51	U	SMPL
K1760-12 - 1	Z2-10	-14	0.0450	-7470A	PPB		08 Mar 2019 15:54:57	U	SMPL
K1760-16 - 1	Z2-11	40	0.0575	-7470A	PPB		08 Mar 2019 15:57:05	U	SMPL
K1760-20 - 1	Z2-12	24	0.0538	-7470A	PPB		08 Mar 2019 15:59:12	U	SMPL
K1780-04 - 1	SOIL-A	21	0.0531	-7470A	PPB		08 Mar 2019 16:01:20	U	SMPL
K1780-08 - 1	SOIL-B	22	0.0533	-7470A	PPB		08 Mar 2019 16:03:28	U	SMPL
PB117744BL - 1	PBW	25	0.0540	-7470A	PPB		08 Mar 2019 16:05:34	U	SMPL
CCV23 - 1	CCV23	20962	4.8965	-7470A	PPB		08 Mar 2019 16:07:41	U	SMPL
CCB23 - 1	CCB23	-32	0.0408	-7470A	PPB		08 Mar 2019 16:09:48	U	SMPL
PB117744BS - 1	LCSW	17740	4.1513	-7470A	PPB		08 Mar 2019 16:11:55	U	SMPL
K1761-02 - 1	N4871	-20	0.0436	-7470A	PPB		08 Mar 2019 16:14:02	U	SMPL
K1761-02DUP - 1	N4871DUP	-4	0.0473	-7470A	PPB		08 Mar 2019 16:16:09	U	SMPL
K1761-02MS - 1	N4871MS	16756	3.9237	-7470A	PPB		08 Mar 2019 16:18:16	U	SMPL
K1761-02MSD - 1	N4871MSD	16753	3.9230	-7470A	PPB		08 Mar 2019 16:20:23	U	SMPL
PB117745BL - 1	PBW	247	0.1054	-7470A	PPB		08 Mar 2019 16:22:30	U	SMPL
PB117745BS - 1	LCSW	17789	4.1626	-7470A	PPB		08 Mar 2019 16:24:37	U	SMPL
K1783-01 - 1	SW-4	0	0.0482	-7470A	PPB		08 Mar 2019 16:26:45	U	SMPL
PB117745BL - 1	PBW	1	0.0485	-7470A	PPB		08 Mar 2019 16:28:09	U	SMPL
PB117745BS - 1	LCSW	17494	4.0944	-7470A	PPB		08 Mar 2019 16:30:16	U	SMPL
K1783-01 - 1	SW-4	256	0.1074	-7470A	PPB		08 Mar 2019 16:32:23	U	SMPL
K1783-02 - 1	SW-3	372	0.1343	-7470A	PPB		08 Mar 2019 16:34:32	U	SMPL
K1783-03 - 1	2019-03-05-DUP-SW	121	0.0762	-7470A	PPB		08 Mar 2019 16:36:38	U	SMPL
CCV24 - 1	CCV24	20378	4.7614	-7470A	PPB		08 Mar 2019 16:38:45	U	SMPL
CCB24 - 1	CCB24	-43	0.0383	-7470A	PPB		08 Mar 2019 16:40:51	U	SMPL
K1783-04 - 1	SW-2	134	0.0792	-7470A	PPB		08 Mar 2019 16:42:58	U	SMPL
K1783-04DUP - 1	SW-2DUP	146	0.0820	-7470A	PPB		08 Mar 2019 16:45:04	U	SMPL
K1783-05 - 1	K1783-04MS	16456	3.8543	-7470A	PPB		08 Mar 2019 16:47:11	U	SMPL
K1783-06 - 1	K1783-04MSD	16568	3.8802	-7470A	PPB		08 Mar 2019 16:49:18	U	SMPL
K1783-07 - 1	SW-1	288	0.1148	-7470A	PPB		08 Mar 2019 16:51:24	U	SMPL
K1783-08 - 1	FB-2019-03-05	-10	0.0459	-7470A	PPB		08 Mar 2019 16:57:02	U	SMPL
PB117746BL - 1	PBW	69	0.0642	-7470A	PPB		08 Mar 2019 16:59:09	U	SMPL
PB117746BS - 1	LCSW	16148	3.7831	-7470A	PPB		08 Mar 2019 17:01:17	U	SMPL
K1788-01 - 1	TS-4	192	0.0926	-7470A	PPB		08 Mar 2019 17:03:25	U	SMPL
K1788-01DUP - 1	TS-4DUP	202	0.0950	-7470A	PPB		08 Mar 2019 17:05:32	U	SMPL
CCV25 - 1	CCV25	20199	4.7200	-7470A	PPB		08 Mar 2019 17:07:38	U	SMPL
CCB25 - 1	CCB25	-73	0.0313	-7470A	PPB		08 Mar 2019 17:09:45	U	SMPL
K1788-01MS - 1	TS-4MS	16938	3.9658	-7470A	PPB		08 Mar 2019 17:11:51	U	SMPL
K1788-01MSD - 1	TS-4MSD	16421	3.8462	-7470A	PPB		08 Mar 2019 17:13:58	U	SMPL
K1788-02 - 1	TS-6	133	0.0790	-7470A	PPB		08 Mar 2019 17:16:05	U	SMPL
K1788-03 - 1	TS-12	260	0.1084	-7470A	PPB		08 Mar 2019 17:18:11	U	SMPL
PB117759BL - 1	PBW	-22	0.0431	-7470A	PPB		08 Mar 2019 17:20:18	U	SMPL
PB117759BS - 1	LCSW	17386	4.0694	-7470A	PPB		08 Mar 2019 17:22:25	U	SMPL
K1693-02 - 1	OR-01-030719-A	35	0.0563	-7470A	PPB		08 Mar 2019 17:24:33	U	SMPL
K1693-02DUP - 1	OR-01-030719-ADUP	60	0.0621	-7470A	PPB		08 Mar 2019 17:26:41	U	SMPL
K1693-02MS - 1	OR-01-030719-AMS	16552	3.8765	-7470A	PPB		08 Mar 2019 17:28:48	U	SMPL
K1693-02MSD - 1	OR-01-030719-AMSD	16444	3.8515	-7470A	PPB		08 Mar 2019 17:30:56	U	SMPL
CCV26 - 1	CCV26	20630	4.8197	-7470A	PPB		08 Mar 2019 17:33:03	U	SMPL
CCB26 - 1	CCB26	-38	0.0394	-7470A	PPB		08 Mar 2019 17:35:10	U	SMPL
K1693-04 - 1	OR-01-030719-B	62	0.0626	-7470A	PPB		08 Mar 2019 17:37:17	U	SMPL
K1787-04 - 1	SP-1	71	0.0647	-7470A	PPB		08 Mar 2019 17:39:24	U	SMPL
K1797-01 - 1	3-7-A	58	0.0616	-7470A	PPB		08 Mar 2019 17:41:31	U	SMPL
K1797-02 - 1	3-7-B	86	0.0681	-7470A	PPB		08 Mar 2019 17:43:37	U	SMPL
PB117743TB - 1	PB117769TB	38	0.0570	-7470A	PPB		08 Mar 2019 17:45:45	U	SMPL
PB117744TB - 1	PB117681TB	34	0.0561	-7470A	PPB		08 Mar 2019 17:47:52	U	SMPL
PB117746TB - 1	PB117717TB	61	0.0623	-7470A	PPB		08 Mar 2019 17:49:59	U	SMPL
PB117759TB - 1	PB117724TB	112	0.0741	-7470A	PPB		08 Mar 2019 17:52:07	U	SMPL

LB101266 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Mar 2019 14:25:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
K1757-04LX5 - 1		2	0.0487		-7470A	PPB	08 Mar 2019 17:54:14	U	SMPL
K1757-04A - 1		14797	3.4706		-7470A	PPB	08 Mar 2019 17:56:21	U	SMPL
CCV27 - 1	CCV27	19655	4.5942		-7470A	PPB	08 Mar 2019 17:58:29	U	SMPL
CCB27 - 1	CCB27	-70	0.0320		-7470A	PPB	08 Mar 2019 18:00:36	U	SMPL
K1761-02LX5 - 1		-11	0.0457		-7470A	PPB	08 Mar 2019 18:02:44	U	SMPL
K1761-02A - 1		15819	3.7070		-7470A	PPB	08 Mar 2019 18:04:51	U	SMPL
K1783-04LX5 - 1		3	0.0489		-7470A	PPB	08 Mar 2019 18:06:58	U	SMPL
K1783-04A - 1		16195	3.7939		-7470A	PPB	08 Mar 2019 18:09:05	U	SMPL
K1788-01LX5 - 1		27	0.0545		-7470A	PPB	08 Mar 2019 18:11:12	U	SMPL
K1788-01A - 1		16337	3.8268		-7470A	PPB	08 Mar 2019 18:13:20	U	SMPL
K1693-02LX5 - 1		-49	0.0369		-7470A	PPB	08 Mar 2019 18:15:27	U	SMPL
K1693-02A - 1		16275	3.8124		-7470A	PPB	08 Mar 2019 18:17:35	U	SMPL
CCV28 - 1	CCV28	20538	4.7984		-7470A	PPB	08 Mar 2019 18:19:44	U	SMPL
CCB28 - 1	CCB28	-79	0.0300		-7470A	PPB	08 Mar 2019 18:21:51	U	SMPL