



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101288

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	03/09/2019	09:10
2	K1834-18		1	7	>100.0	>212.0	>212.0	03/09/2019	09:40
3	K1834-18DUP		1	7	>100.0	>212.0	>212.0	03/09/2019	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 101288

WorkList Name : FP3919

WorkList ID : 123213

Date : 03-09-2019 08:33:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2019	Water	K1834-18	Flash Point	Cool 4 deg C	PSEG01	K31	COMP-2-AU	03/08/2019	1010A

Date/Time 03/09/19 08:40
 Received by: KS
 Relinquished by: KS

Date/Time 03/09/19 10:20
 Received by: KS
 Relinquished by: KS