



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101291

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	03/11/2019	12:40
2	K1834-22		1	8	>100.0	>212.0	>212.0	03/11/2019	12:58
3	K1834-22DUP		1	8	>100.0	>212.0	>212.0	03/11/2019	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB101291

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP31119-1

WorkList ID : 123259

Date : 03-11-2019 12:11:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2019	Solid	K1834-22	C Flash Point	Cool 4 deg C	PSEG01	K31	23187	03/08/2019	1010A

Date/Time 03/11/19 12:15
Received by: KS
Relinquished by: CP

Date/Time 03/11/19 13:20
Received by: CP
Relinquished by: KS