

LB101293


245.1

INSTRUMENT ID : CV1

Linear

 μ Abs.:

43304



A= 0.0000e+000

B= 2.3306e-004

C= -5.6206e-002

Rho= 0.9997926

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.069	-0.069	-55	0.000	-55				
0.05	0.050	0.059	0.009	494	0.0 %	494				
0.2	0.200	0.217	0.017	1171	0.0 %	1171				
2.5	2.500	2.536	0.036	11123	0.0 %	11123				
5.0	5.000	5.112	0.112	22174	0.0 %	22174				
7.5	7.500	7.359	-0.141	31818	0.0 %	31818				
10.0	10.000	10.036	0.036	43304	0.0 %	43304				

LB101293 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Mar 2019 14:54:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0-0	-55	-	0.0000	245.1	PPB	11 Mar 2019 15:44:02	S	Std
0.05 - 1	S0-05	494	-	0.0500	245.1	PPB	11 Mar 2019 15:46:19	S	Std
0.2 - 1	S0-2	1171	-	0.2000	245.1	PPB	11 Mar 2019 15:48:36	S	Std
2.5 - 1	S2-5	11123	-	2.5000	245.1	PPB	11 Mar 2019 15:50:53	S	Std
5.0 - 1	S5-0	22174	-	5.0000	245.1	PPB	11 Mar 2019 15:53:10	S	Std
7.5 - 1	S7-5	31818	-	7.5000	245.1	PPB	11 Mar 2019 15:55:26	S	Std
10.0 - 1	S10-0	43304	-	10.0000	245.1	PPB	11 Mar 2019 15:57:42	S	Std
ICV49 - 1	ICV49	17363	3.9904	-245.1		PPB	11 Mar 2019 16:01:54	U	SMPL
ICB49 - 1	ICB49	232	-0.0021	-245.1		PPB	11 Mar 2019 16:04:09	U	SMPL
CCV29 - 1	CCV29	22147	5.1054	-245.1		PPB	11 Mar 2019 16:06:24	U	SMPL
CCB29 - 1	CCB29	232	-0.0021	-245.1		PPB	11 Mar 2019 16:08:39	U	SMPL
CRA - 1	CRA	484	0.0566	-245.1		PPB	11 Mar 2019 16:10:55	U	SMPL
HighStd - 1	HighStd	44504	10.3159	-245.1		PPB	11 Mar 2019 16:13:10	U	SMPL
ChkStd - 1	ChkStd	29507	6.8207	-245.1		PPB	11 Mar 2019 16:30:22	U	SMPL
PB117811BL - 1	PBW	105	-0.0317	-245.1		PPB	11 Mar 2019 16:32:39	U	SMPL
PB117811BS - 1	LCSW	18104	4.1631	-245.1		PPB	11 Mar 2019 16:34:55	U	SMPL
K1756-02 - 1	COMP	344	0.0240	-245.1		PPB	11 Mar 2019 16:37:12	U	SMPL
K1756-02DUP - 1	COMPDUP	335	0.0219	-245.1		PPB	11 Mar 2019 16:39:28	U	SMPL
K1756-02MS - 1	COMPMS	17122	3.9343	-245.1		PPB	11 Mar 2019 16:41:45	U	SMPL
K1756-02MSD - 1	COMPMSD	17268	3.9683	-245.1		PPB	11 Mar 2019 16:44:02	U	SMPL
K1798-01 - 1	15220032-1	196	-0.0105	-245.1		PPB	11 Mar 2019 16:46:18	U	SMPL
CCV30 - 1	CCV30	22230	5.1247	-245.1		PPB	11 Mar 2019 16:48:33	U	SMPL
CCB30 - 1	CCB30	239	-0.0005	-245.1		PPB	11 Mar 2019 16:50:48	U	SMPL
K1799-01 - 1	32220008-3	142	-0.0231	-245.1		PPB	11 Mar 2019 16:53:04	U	SMPL
K1756-02LX5 - 1		64	-0.0413	-245.1		PPB	11 Mar 2019 16:55:20	U	SMPL
K1756-02A - 1		17069	3.9219	-245.1		PPB	11 Mar 2019 16:57:35	U	SMPL
CCV31 - 1	CCV31	21677	4.9959	-245.1		PPB	11 Mar 2019 16:59:51	U	SMPL
CCB31 - 1	CCB31	206	-0.0082	-245.1		PPB	11 Mar 2019 17:02:07	U	SMPL