



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By :

Run Number: LB101304

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/12/2019	14:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/12/2019	14:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/12/2019	14:02
4	ICV	1	Water	NA	NA	20.2	6.99	03/12/2019	14:03
5	CCV1	1	Water	NA	NA	20.2	2.01	03/12/2019	14:05
6	K1858-01	1	Water	NA	NA	22.2	4.94	03/12/2019	14:06
7	K1858-01DUP	1	Water	NA	NA	22.3	4.95	03/12/2019	14:08
8	CCV2	1	Water	NA	NA	20.2	12.01	03/12/2019	14:10

18101304

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k1858

WorkList ID : 123315

Date : 03-12-2019 13:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/11/2019	Water	K1858-01 - 17	pH	Cool 4 deg C	STER08	P11	POOL-WATER	03/11/2019	SM 4500-H B

Date/Time 03-12-19 16:00
 Received by: Jg
 Relinquished by: CP

Date/Time 03-12-19 16:00
 Received by: Jg
 Relinquished by: CP