

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/12/2019 16:02
=====

Test: Phenolics-

Sample Id	Result ^{mg/d}	Dil. 1 +	Response	Errors
ICV1	1.009	0.0	0.120	
ICB1	0.000	0.0	0.017	
CCV1	1.024	0.0	0.122	
CCB1	-0.002	0.0	0.017	
PB117815BL	0.001	0.0	0.017	
PB117815BS	1.033	0.0	0.123	
K1833-27	2.735	0.0	0.298	Test limit high
CCV2	1.016	0.0	0.121	
CCB2	0.001	0.0	0.017	
K1833-27DLX2	1.452	0.0	0.166	
CCV3	1.016	0.0	0.121	
CCB3	-0.001	0.0	0.017	

N 12
Mean 0.774
SD 0.8314
CV% 107.45

Aquakem v. 7.2AQ1

Results from time period:

Tue Mar 12 15:08:53 2019

Tue Mar 12 16:00:25 2019

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0476	mg/l	3/12/2019 15:08:53	
0.1PPM	A	Phenolics-	P	0.0957	mg/l	3/12/2019 15:08:54	
0.25PPM	A	Phenolics-	P	0.2461	mg/l	3/12/2019 15:08:55	
0.50PPM	A	Phenolics-	P	0.5019	mg/l	3/12/2019 15:08:56	
1.0PPM	A	Phenolics-	P	1.017	mg/l	3/12/2019 15:08:57	
2.0PPM	A	Phenolics-	P	1.9918	mg/l	3/12/2019 15:08:58	
ICV1	S	Phenolics-	P	1.0091	mg/l	3/12/2019 15:31:46	
ICB1	S	Phenolics-	P	0.0003	mg/l	3/12/2019 15:31:47	
CCV1	S	Phenolics-	P	1.0243	mg/l	3/12/2019 15:31:48	
CCB1	S	Phenolics-	P	-0.0019	mg/l	3/12/2019 15:31:49	
PB117815BL	S	Phenolics-	P	0.0011	mg/l	3/12/2019 15:31:50	
PB117815BS	S	Phenolics-	P	1.0334	mg/l	3/12/2019 15:31:51	
K1833-27	S	Phenolics-	P	2.7348	mg/l	3/12/2019 15:31:52	
CCV2	S	Phenolics-	P	1.016	mg/l	3/12/2019 15:31:56	
CCB2	S	Phenolics-	P	0.001	mg/l	3/12/2019 15:31:57	
K1833-27DLX2	S	Phenolics-	P	1.4522	mg/l	3/12/2019 16:00:21	
CCV3	S	Phenolics-	P	1.016	mg/l	3/12/2019 16:00:22	
CCB3	S	Phenolics-	P	-0.0013	mg/l	3/12/2019 16:00:23	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/12/2019 15:09


Test Phenolics-

Accepted 3/12/2019 15:09

Factor 9.732
Bias 0.017

Coeff. of det. 0.999857

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.022	0.0476	0.0500	
2	Phenol-2	0.027	0.0957	0.1000	
3	Phenol-2	0.042	0.2461	0.2500	
4	Phenol-2	0.068	0.5019	0.5000	
5	Phenol-2	0.121	1.0170	1.0000	
6	Phenol-2	0.221	1.9918	2.0000	