

dB101314

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AS

Instrument ID : Konelab

3/12/2019 16:27

Test: Phenolics-

mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.006	0.0	0.120	
ICB1	-0.003	0.0	0.016	
CCV1	1.016	0.0	0.121	
CCB1	-0.004	0.0	0.016	
PB117816BL	-0.002	0.0	0.016	
PB117816BS	1.011	0.0	0.121	
K1833-26	2.933	0.0	0.318	Test limit high
K1833-26DLX2	1.468	0.0	0.168	
CCV2	1.018	0.0	0.121	
CCB2	-0.003	0.0	0.016	

N 10
Mean 0.844
SD 0.9271
CV% 109.85

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/12/2019 16:27


Test Phenolics-

Accepted 3/12/2019 15:09

Factor 9.732
Bias 0.017

Coeff. of det. 0.999857

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.022	0.0476	0.0500	
2	Phenol-2	0.027	0.0957	0.1000	
3	Phenol-2	0.042	0.2461	0.2500	
4	Phenol-2	0.068	0.5019	0.5000	
5	Phenol-2	0.121	1.0170	1.0000	
6	Phenol-2	0.221	1.9918	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Mar 12 16:27:11 2019

Tue Mar 12 16:27:20 2019

✓ 42

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0476	mg/l	3/12/2019 15:08:53	
0.1PPM	A	Phenolics-	P	0.0957	mg/l	3/12/2019 15:08:54	
0.25PPM	A	Phenolics-	P	0.2461	mg/l	3/12/2019 15:08:55	
0.50PPM	A	Phenolics-	P	0.5019	mg/l	3/12/2019 15:08:56	
1.0PPM	A	Phenolics-	P	1.017	mg/l	3/12/2019 15:08:57	
2.0PPM	A	Phenolics-	P	1.9918	mg/l	3/12/2019 15:08:58	
ICV1	S	Phenolics-	P	1.0056	mg/l	3/12/2019 16:27:11	
ICB1	S	Phenolics-	P	-0.0029	mg/l	3/12/2019 16:27:12	
CCV1	S	Phenolics-	P	1.016	mg/l	3/12/2019 16:27:13	
CCB1	S	Phenolics-	P	-0.0035	mg/l	3/12/2019 16:27:14	
PB117816BL	S	Phenolics-	P	-0.0022	mg/l	3/12/2019 16:27:15	
PB117816BS	S	Phenolics-	P	1.0109	mg/l	3/12/2019 16:27:16	
K1833-26	S	Phenolics-	P	2.9327	mg/l	3/12/2019 16:27:17	
K1833-26DLX2	S	Phenolics-	P	1.4682	mg/l	3/12/2019 16:27:18	
CCV2	S	Phenolics-	P	1.0182	mg/l	3/12/2019 16:27:19	
CCB2	S	Phenolics-	P	-0.0031	mg/l	3/12/2019 16:27:20	