



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101317

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	03/12/2019	15:20
2	K1844-10		1	9	>100.0	>212.0	>212.0	03/12/2019	15:45
3	K1844-10DUP		1	9	>100.0	>212.0	>212.0	03/12/2019	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 101317

WorkList Name : FP31219

WorkList ID : 123313

Date : 03-12-2019 13:31:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/15/2019	Solid	K1844-10	Flash Point	Cool 4 deg C	PSEG01	K31	TOTE-1-3	03/11/2019	1010A

Date/Time 03/12/19 14:50
 Received by: KS
 Relinquished by: CP

Date/Time 03/12/19 16:15
 Received by: CP
 Relinquished by: KS