



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB101332

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP72523
Turbidity std, 40NTU	WP72515
turbidity std 5 NTU	WP72518
turbidity std 20 NTU	WP72517
Turbidity std 1 NTU	WP72519
turbidity std 0 NTU	WP72512
turbidity std 10 NTU	WP72520
TURBIDITY LOD STD, 0.5NTU	WP72527

Intercept: 0.0497

Slope: 0.9593

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.250	03/13/2019	13:10
2	CAL2	1	1	0.995	03/13/2019	13:15
3	CAL3	5	1	4.870	03/13/2019	13:21
4	CAL4	10	1	9.790	03/13/2019	13:26
5	CAL5	20	1	18.600	03/13/2019	13:32
6	CAL6	40	1	38.700	03/13/2019	13:37



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB101332

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.000	10.372	03/13/2019	13:42
2	ICB		1	0.262	0.221	03/13/2019	13:47
3	CCV1	10	1	9.980	10.352	03/13/2019	13:50
4	CCB1		1	0.255	0.214	03/13/2019	13:54
5	LB101332BL		1	0.261	0.220	03/13/2019	13:58
6	K1560-12		1	0.707	0.685	03/13/2019	14:02
7	K1887-01		1	0.546	0.517	03/13/2019	14:05
8	K1887-02		1	6.020	6.224	03/13/2019	14:09
9	K1887-03		1	0.545	0.516	03/13/2019	14:13
10	K1887-04		1	0.591	0.564	03/13/2019	14:17
11	K1887-04DUP		1	0.589	0.562	03/13/2019	14:20
12	CCV2	10	1	9.970	10.341	03/13/2019	14:22
13	CCB2		1	0.258	0.217	03/13/2019	14:26

LB 101332

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity031319

WorkList ID : 123366

Date : 03-13-2019 11:52:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/23/2019	Water	K1560-12	Turbidity	Cool 4 deg C	CHEM02	Wet Chem L	MDL-MDL-WATER-06-QT1-2C	02/21/2019	SM2130 B
03/14/2019	Water	K1887-01	Turbidity	Cool 4 deg C	LOCK01	P41	MW-1	03/12/2019	SM2130 B
03/14/2019	Water	K1887-02	Turbidity	Cool 4 deg C	LOCK01	P41	MW-2	03/12/2019	SM2130 B
03/14/2019	Water	K1887-03	Turbidity	Cool 4 deg C	LOCK01	P41	MW-3	03/12/2019	SM2130 B
03/14/2019	Water	K1887-04	Turbidity	Cool 4 deg C	LOCK01	P41	MW-4	03/12/2019	SM2130 B

Date/Time 03-13-19 19:29
Received by: AK
Relinquished by: CP

Date/Time 03-13-19 14:30
Received by: CP
Relinquished by: AK