



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB101340
Analysis Date: 03/13/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 03/13/2019
Extraction IN Time: 07:25
Extraction OUT Time: 08:10

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB101340BL	LB101340BL	WATER	1.3	1000	100	2.8413	2.8413	3.02	2.8413	2.8413	0.0000	0
2	LB101340BS	LB101340BS	WATER	1.3	1000	100	3.1220	3.1220	3.01	3.1407	3.1407	0.0187	18.7
3	LB101340BSD	LB101340BSD	WATER	1.3	1000	100	3.0589	3.0589	3.03	3.0778	3.0778	0.0189	18.9
4	K1833-18	RR-TPH1L-WP	WATER	1.3	880	100	3.1033	3.1033	3.04	3.1285	3.1285	0.0252	28.64
5	K1844-02	DRUM-1-4	WATER	1.6	1000	100	3.2371	3.2371	3.05	3.2403	3.2403	0.0032	3.2



QC Batch# LB101340 Test: TPH Analysis Date: 03/13/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2494
pH Paper 0-14	M4240
Sodium Sulfate	EP1930
1:1 HCL	WP69312
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP71442
LCSWD	10.00 ML	WP71443
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:31</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:30</u>		
Bal Check Time:	<u>07:40</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:05</u>
		Out Time:	<u>09:30</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:11</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>10:35</u>		
Bal Check Time:	<u>12:05</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:35</u>
		Out Time:	<u>11:10</u>		

LB101340

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K1844

WorkList ID : 123363

Date : 03-13-2019 7:15:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/15/2019	Water	K1844-02-12	TPH	Conc H2SO4 to	PSEG01	K31	DRUM-1-4	03/11/2019	1664A

Date/Time 03-13-19 7:20
Received by: JD
Relinquished by: SD

Date/Time 03-13-19 09:25
Received by: SD
Relinquished by: JD