

2B101343

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

3/13/2019 14:19

Test: Total CN

Sample Id	Result <i>ug/d</i>	Dil. 1 +	Response	Errors
ICV1	97.991	0.0	0.093	
ICB1	0.540	0.0	0.005	
CCV1	243.546	0.0	0.225	
CCB1	0.485	0.0	0.004	
PB117860BL	0.492	0.0	0.004	
PB117860BS	98.778	0.0	0.094	
MIDPB117860	243.832	0.0	0.225	
K1833-23	903.340	0.0	0.824	Test limit high
K1833-23DLX5	169.914	0.0	0.158	
CCV2	255.333	0.0	0.236	
CCB2	0.492	0.0	0.004	

N	11
Mean	183.159
SD	260.6149
CV%	142.29

Aquakem v. 7.2AQ1

AS

Results from time period:

Wed Mar 13 14:13:50 2019

Wed Mar 13 14:17:14 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.3651	µg/l	3/13/2019 10:11:09	
5.OPPBCN	A	Total CN	P	4.7637	µg/l	3/13/2019 10:11:10	
10PPBCN	A	Total CN	P	9.6897	µg/l	3/13/2019 10:11:11	
50PPBCN	A	Total CN	P	50.2264	µg/l	3/13/2019 10:11:12	
100PPBCN	A	Total CN	P	100.1116	µg/l	3/13/2019 10:11:13	
250PPBCN	A	Total CN	P	249.7596	µg/l	3/13/2019 10:11:14	
500PPBCN	A	Total CN	P	500.0838	µg/l	3/13/2019 10:11:15	
ICV1	S	Total CN	P	97.991	µg/l	3/13/2019 14:13:50	
ICB1	S	Total CN	P	0.5398	µg/l	3/13/2019 14:13:51	
CCV1	S	Total CN	P	243.5462	µg/l	3/13/2019 14:13:52	
CCB1	S	Total CN	P	0.4854	µg/l	3/13/2019 14:13:53	
PB117860BL	S	Total CN	P	0.4921	µg/l	3/13/2019 14:13:54	
PB117860BS	S	Total CN	P	98.7783	µg/l	3/13/2019 14:13:55	
MIDPB117860	S	Total CN	P	243.8317	µg/l	3/13/2019 14:13:56	
K1833-23	S	Total CN	P	903.3405	µg/l	3/13/2019 14:13:57	
K1833-23DLX5	S	Total CN	P	169.9139	µg/l	3/13/2019 14:14:00	
CCV2	S	Total CN	P	255.3328	µg/l	3/13/2019 14:17:13	
CCB2	S	Total CN	P	0.492	µg/l	3/13/2019 14:17:14	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

Instrument ID : Konelab

3/13/2019 14:22


Test Total CN

Accepted 3/13/2019 10:11

Factor 1102
 Bias 0.004

Coeff. of det. 0.999998

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.3651	0.0000	
2	5.0PPBCN	0.008	4.7637	5.0000	
3	10PPBCN	0.013	9.6897	10.0000	
4	50PPBCN	0.050	50.2264	50.0000	
5	100PPBCN	0.095	100.1116	100.0000	
6	250PPBCN	0.231	249.7596	250.0000	
7	500PPBCN	0.458	500.0838	500.0000	