



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By :

Run Number: LB101344

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/13/2019	14:35
2	CAL2	1	Water	NA	NA	20.4	7.01	03/13/2019	14:36
3	CAL3	1	Water	NA	NA	20.4	10.02	03/13/2019	14:37
4	ICV	1	Water	NA	NA	20.4	7.00	03/13/2019	14:39
5	CCV1	1	Water	NA	NA	20.3	2.01	03/13/2019	14:40
6	K1887-01	1	Water	NA	NA	14.7	5.84	03/13/2019	14:41
7	K1887-01DUP	1	Water	NA	NA	14.8	5.85	03/13/2019	14:43
8	K1887-02	1	Water	NA	NA	15.2	6.46	03/13/2019	14:45
9	K1887-03	1	Water	NA	NA	17.4	6.46	03/13/2019	14:46
10	K1887-04	1	Water	NA	NA	17.0	6.50	03/13/2019	14:48
11	CCV2	1	Water	NA	NA	20.3	12.01	03/13/2019	14:50

LB101344

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1887

WorkList ID : 123378

Date : 03-13-2019 14:28:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2019	Water	K1887-01-01	pH	Cool 4 deg C	LOCK01	P41	MW-1	03/12/2019	SM 4500-H B
03/12/2019	Water	K1887-02-01	pH	Cool 4 deg C	LOCK01	P41	MW-2	03/12/2019	SM 4500-H B
03/12/2019	Water	K1887-03-01	pH	Cool 4 deg C	LOCK01	P41	MW-3	03/12/2019	SM 4500-H B
03/12/2019	Water	K1887-04-01	pH	Cool 4 deg C	LOCK01	P41	MW-4	03/12/2019	SM 4500-H B

Date/Time 03-13-19 14:30
 Received by: JD
 Relinquished by: CA

Date/Time 03-13-19 15:15
 Received by: CA
 Relinquished by: JD