



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101361

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/13/2019	17:25
2	CAL2	1	Water	NA	NA	20.2	7.01	03/13/2019	17:26
3	CAL3	1	Water	NA	NA	20.2	10.03	03/13/2019	17:27
4	ICV	1	Water	NA	NA	20.2	6.99	03/13/2019	17:30
5	CCV1	1	Water	NA	NA	20.2	2.02	03/13/2019	17:40
6	K1902-11	1	Solid	20.02	20	21.6	6.87	03/13/2019	17:41
7	K1902-11DUP	1	Solid	20.03	20	21.6	6.88	03/13/2019	17:42
8	CCV2	1	Water	NA	NA	20.1	12.01	03/13/2019	17:44

28101361

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K1902

WorkList ID : 123384

Date : 03-13-2019 17:15:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/13/2019	Solid	K1902-11	pH	Cool 4 deg C	PSEG01	K31	30245	03/13/2019	9041A

Date/Time 03-13-19 17:00
Received by: gn
Relinquished by: cn

Date/Time 03-13-19 18:30
Received by: JP
Relinquished by: cn