



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101362

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/13/2019	17:25
2	CAL2	1	Water	NA	NA	20.2	7.01	03/13/2019	17:26
3	CAL3	1	Water	NA	NA	20.2	10.03	03/13/2019	17:27
4	ICV	1	Water	NA	NA	20.2	6.99	03/13/2019	17:30
5	CCV1	1	Water	NA	NA	20.2	2.01	03/13/2019	17:31
6	K1892-01	1	Solid	20.02	20	21.5	11.35	03/13/2019	17:32
7	K1895-02	1	Solid	20.04	20	21.0	7.46	03/13/2019	17:34
8	K1897-04	1	Solid	20.03	20	21.3	5.84	03/13/2019	17:35
9	K1897-09	1	Solid	20.04	20	21.1	5.40	03/13/2019	17:36
10	K1897-09DUP	1	Solid	20.03	20	21.2	5.41	03/13/2019	17:37
11	CCV2	1	Water	NA	NA	20.1	12.01	03/13/2019	17:38

18101362

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K1897

WorkList ID : 123385

Date : 03-13-2019 17:15:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/13/2019	Solid	K1892-01	Corrosivity	Cool 4 deg C	PSEG01	K31	GARAGE-FLOOR	03/13/2019	9045D
03/13/2019	Solid	K1895-02	Corrosivity	Cool 4 deg C	PSEG01	K31	OILY-SOIL	03/13/2019	9045D
03/13/2019	Solid	K1897-04	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-1	03/13/2019	9045D
03/13/2019	Solid	K1897-09	Corrosivity	Cool 4 deg C	PSEG01	K31	TP-2	03/13/2019	9045D

Date/Time 03.13.19 17:15:49
 Received by: J
 Relinquished by: M

Date/Time 03.13.19 18:36
 Received by: J
 Relinquished by: M