

LB101366


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46273



A= 0.0000e+000

B= 2.1620e-004

C= 5.9036e-002

Rho= 0.9998970

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.043	0.043	-75	0.000	-75				
0.2	0.200	0.227	0.027	779	0.0 %	779				
2.5	2.500	2.480	-0.020	11196	0.0 %	11196				
5.0	5.000	4.902	-0.098	22401	0.0 %	22401				
7.5	7.500	7.485	-0.015	34345	0.0 %	34345				
10.0	10.000	10.063	0.063	46273	0.0 %	46273				

LB101366 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Mar 2019 10:22:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-75	-	0.0000	7470A	PPB	14 Mar 2019 10:43:42	S	Std
0.2 - 1	50.2	779	-	0.2000	7470A	PPB	14 Mar 2019 10:45:50	S	Std
2.5 - 1	52.5	11196	-	2.5000	7470A	PPB	14 Mar 2019 10:47:58	S	Std
5.0 - 1	55.0	22401	-	5.0000	7470A	PPB	14 Mar 2019 10:50:06	S	Std
7.5 - 1	57.5	34345	-	7.5000	7470A	PPB	14 Mar 2019 10:52:14	S	Std
10.0 - 1	510.0	46273	-	10.0000	7470A	PPB	14 Mar 2019 10:54:21	S	Std
ICV54 - 1	ICV54	19566	4.2893	-	7470A	PPB	14 Mar 2019 11:02:04	U	SMPL
ICB54 - 1	ICB54	-58	0.0465	-	7470A	PPB	14 Mar 2019 11:04:10	U	SMPL
CCV61 - 1	CCV61	22521	4.9281	-	7470A	PPB	14 Mar 2019 11:06:15	U	SMPL
CCB61 - 1	CCB61	-19	0.0549	-	7470A	PPB	14 Mar 2019 11:08:21	U	SMPL
CRA - 1	CRA	769	0.2253	-	7470A	PPB	14 Mar 2019 11:10:27	U	SMPL
HighStd - 1	HighStd	46884	10.1955	-	7470A	PPB	14 Mar 2019 11:12:33	U	SMPL
ChkStd - 1	ChkStd	32310	7.0445	-	7470A	PPB	14 Mar 2019 11:14:39	U	SMPL
PB117913BL - 1	PBW	-64	0.0452	-	7470A	PPB	14 Mar 2019 11:16:47	U	SMPL
PB117913BS - 1	LCSW	17023	3.7395	-	7470A	PPB	14 Mar 2019 11:18:54	U	SMPL
K1693-06 - 1	OR-02-031219-A	-16	0.0556	-	7470A	PPB	14 Mar 2019 11:21:01	U	SMPL
K1693-06DUP - 1	OR-02-031219-ADUP	23	0.0640	-	7470A	PPB	14 Mar 2019 11:23:08	U	SMPL
K1693-06MS - 1	OR-02-031219-AMS	17001	3.7347	-	7470A	PPB	14 Mar 2019 11:25:16	U	SMPL
K1693-06MSD - 1	OR-02-031219-AMSD	16927	3.7187	-	7470A	PPB	14 Mar 2019 11:27:24	U	SMPL
K1693-08 - 1	OR-02-031219-B	-11	0.0567	-	7470A	PPB	14 Mar 2019 11:29:30	U	SMPL
CCV62 - 1	CCV62	22714	4.9699	-	7470A	PPB	14 Mar 2019 11:31:36	U	SMPL
CCB62 - 1	CCB62	-22	0.0543	-	7470A	PPB	14 Mar 2019 11:33:42	U	SMPL
K1847-19 - 1	SUP-7B-C-1-031119	36	0.0668	-	7470A	PPB	14 Mar 2019 11:35:49	U	SMPL
K1847-20 - 1	SUP-7B-C-2-031119	64	0.0729	-	7470A	PPB	14 Mar 2019 11:37:55	U	SMPL
K1847-21 - 1	SUP-7B-C-3-031119	43	0.0683	-	7470A	PPB	14 Mar 2019 11:40:01	U	SMPL
K1847-22 - 1	SUP-7B-C-4-031119	13	0.0618	-	7470A	PPB	14 Mar 2019 11:42:08	U	SMPL
K1847-23 - 1	SUP-7B-C-5-031119	5	0.0601	-	7470A	PPB	14 Mar 2019 11:44:15	U	SMPL
K1863-01 - 1	MCH-CAULK	45	0.0688	-	7470A	PPB	14 Mar 2019 11:46:23	U	SMPL
K1864-04 - 1	WINDOW-CAULK	18	0.0629	-	7470A	PPB	14 Mar 2019 11:48:30	U	SMPL
K1865-01 - 1	DRUM-1-2	4	0.0599	-	7470A	PPB	14 Mar 2019 11:50:38	U	SMPL
K1869-08 - 1	SP-1	28	0.0651	-	7470A	PPB	14 Mar 2019 11:52:46	U	SMPL
K1869-16 - 1	SP-2	34	0.0664	-	7470A	PPB	14 Mar 2019 11:54:52	U	SMPL
CCV63 - 1	CCV63	23909	5.2282	-	7470A	PPB	14 Mar 2019 11:56:59	U	SMPL
CCB63 - 1	CCB63	-29	0.0528	-	7470A	PPB	14 Mar 2019 11:59:05	U	SMPL
K1870-04 - 1	SP-1	36	0.0668	-	7470A	PPB	14 Mar 2019 12:01:11	U	SMPL
K1870-08 - 1	SP-2	45	0.0688	-	7470A	PPB	14 Mar 2019 12:03:18	U	SMPL
K1870-12 - 1	SP-3	24	0.0642	-	7470A	PPB	14 Mar 2019 12:05:25	U	SMPL
K1870-14 - 1	ROW-CONCRETE-1	19	0.0631	-	7470A	PPB	14 Mar 2019 12:07:31	U	SMPL
K1870-16 - 1	ROW-CONCRETE	22	0.0638	-	7470A	PPB	14 Mar 2019 12:09:38	U	SMPL
K1871-04 - 1	TP-1	43	0.0683	-	7470A	PPB	14 Mar 2019 12:11:45	U	SMPL
K1871-08 - 1	TP-2	35	0.0666	-	7470A	PPB	14 Mar 2019 12:13:52	U	SMPL
K1873-04 - 1	371	20	0.0634	-	7470A	PPB	14 Mar 2019 12:15:59	U	SMPL
PB117914BL - 1	PBW	1	0.0593	-	7470A	PPB	14 Mar 2019 12:18:07	U	SMPL
PB117914BS - 1	LCSW	17513	3.8454	-	7470A	PPB	14 Mar 2019 12:20:13	U	SMPL
CCV64 - 1	CCV64	23614	5.1644	-	7470A	PPB	14 Mar 2019 12:31:37	U	SMPL
CCB64 - 1	CCB64	-28	0.0530	-	7470A	PPB	14 Mar 2019 12:33:43	U	SMPL
K1882-02 - 1	BPS1-TT-MW305D-20190312	2	0.0595	-	7470A	PPB	14 Mar 2019 12:35:49	U	SMPL
K1882-03 - 1	BPS1-TT-MW305S-20190312	11	0.0614	-	7470A	PPB	14 Mar 2019 12:37:56	U	SMPL
K1882-05 - 1	BP-DUP01-20190312	14	0.0621	-	7470A	PPB	14 Mar 2019 12:40:02	U	SMPL
K1882-07 - 1	BPS1-TT-MW304S-20190312	9	0.0610	-	7470A	PPB	14 Mar 2019 12:42:09	U	SMPL
K1882-08 - 1	BPS1-TT-MW304D-20190312	0	0.0590	-	7470A	PPB	14 Mar 2019 12:44:15	U	SMPL
K1882-08DUP - 1	BPS1-TT-MW304D-20190312DUP	16	0.0625	-	7470A	PPB	14 Mar 2019 12:46:22	U	SMPL
K1882-09 - 1	K1882-08MS	17788	3.9048	-	7470A	PPB	14 Mar 2019 12:48:30	U	SMPL
K1882-10 - 1	K1882-08MSD	17830	3.9139	-	7470A	PPB	14 Mar 2019 12:50:38	U	SMPL
K1882-11 - 1	BP-EB01-20190312	-21	0.0545	-	7470A	PPB	14 Mar 2019 12:52:45	U	SMPL
PB117913TB - 1	PB117844TB	2	0.0595	-	7470A	PPB	14 Mar 2019 12:54:52	U	SMPL
CCV65 - 1	CCV65	24656	5.3897	-	7470A	PPB	14 Mar 2019 12:56:59	U	SMPL
CCB65 - 1	CCB65	-40	0.0504	-	7470A	PPB	14 Mar 2019 12:59:05	U	SMPL
K1693-06LX5 - 1		-11	0.0567	-	7470A	PPB	14 Mar 2019 13:01:11	U	SMPL
K1693-06A - 1		17143	3.7654	-	7470A	PPB	14 Mar 2019 13:03:18	U	SMPL
K1882-02LX5 - 1		-17	0.0554	-	7470A	PPB	14 Mar 2019 13:05:24	U	SMPL
K1882-02A - 1		17656	3.8763	-	7470A	PPB	14 Mar 2019 13:07:31	U	SMPL
CCV66 - 1	CCV66	21138	4.6291	-	7470A	PPB	14 Mar 2019 13:09:37	U	SMPL
CCB66 - 1	CCB66	-26	0.0534	-	7470A	PPB	14 Mar 2019 13:11:45	U	SMPL