



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101367

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	03/14/2019	12:35
2	K1902-01		1	9	>100.0	>212.0	>212.0	03/14/2019	12:50
3	K1902-01DUP		1	9	>100.0	>212.0	>212.0	03/14/2019	13:15
4	K1902-02		1	8	>100.0	>212.0	>212.0	03/14/2019	13:35
5	K1902-03		1	7	>100.0	>212.0	>212.0	03/14/2019	13:55
6	K1902-04		1	6	>100.0	>212.0	>212.0	03/14/2019	14:15
7	K1902-05		1	7	>100.0	>212.0	>212.0	03/14/2019	14:40
8	K1902-06		1	7	>100.0	>212.0	>212.0	03/14/2019	14:58
9	K1902-07		1	9	>100.0	>212.0	>212.0	03/14/2019	15:20
10	K1902-08		1	7	>100.0	>212.0	>212.0	03/14/2019	15:45
11	K1902-09		1	8	>100.0	>212.0	>212.0	03/14/2019	16:10
12	K1902-10		1	7	>100.0	>212.0	>212.0	03/14/2019	16:40
13	K1902-12		1	8	>100.0	>212.0	>212.0	03/14/2019	17:02
14	K1902-14		1	9	>100.0	>212.0	>212.0	03/14/2019	17:25
15	K1902-16		1	7	>100.0	>212.0	>212.0	03/14/2019	17:50
16	K1902-18		1	8	>100.0	>212.0	>212.0	03/14/2019	18:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 101367

WorkList Name : FP31419

WorkList ID : 123403

Date : 03-14-2019 09:08:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2019	Water	K1902-01	Flash Point	Cool 4 deg C	PSEG01	K31	51064	03/13/2019	1010A
03/20/2019	Water	K1902-02	Flash Point	Cool 4 deg C	PSEG01	K31	30276	03/13/2019	1010A
03/20/2019	Water	K1902-03	Flash Point	Cool 4 deg C	PSEG01	K31	30298	03/13/2019	1010A
03/20/2019	Water	K1902-04	Flash Point	Cool 4 deg C	PSEG01	K31	30299	03/13/2019	1010A
03/20/2019	Water	K1902-05	Flash Point	Cool 4 deg C	PSEG01	K31	30300	03/13/2019	1010A
03/20/2019	Water	K1902-06	Flash Point	Cool 4 deg C	PSEG01	K31	30301	03/13/2019	1010A
03/20/2019	Water	K1902-07	Flash Point	Cool 4 deg C	PSEG01	K31	30302	03/13/2019	1010A
03/20/2019	Water	K1902-08	Flash Point	Cool 4 deg C	PSEG01	K31	30303	03/13/2019	1010A
03/18/2019	Water	K1902-09	Flash Point	Cool 4 deg C	PSEG01	K31	30282	03/13/2019	1010A
03/20/2019	Water	K1902-10	Flash Point	Cool 4 deg C	PSEG01	K31	51065	03/13/2019	1010A
03/20/2019	Water	K1902-12	Flash Point	Cool 4 deg C	PSEG01	K31	30240-41-42-43-44	03/13/2019	1010A
03/20/2019	Water	K1902-14	Flash Point	Cool 4 deg C	PSEG01	K31	30296-97-30304-05-06	03/13/2019	1010A
03/20/2019	Water	K1902-16	Flash Point	Cool 4 deg C	PSEG01	K31	30307-08-09-10-11	03/13/2019	1010A
03/20/2019	Water	K1902-18	Flash Point	Cool 4 deg C	PSEG01	K31	30312-13-14-15-17	03/13/2019	1010A

Date/Time 03/14/19 12:15
 Received by: KS
 Relinquished by: CP

Date/Time 03/14/19 18:15
 Received by: CP
 Relinquished by: KS